






Article

Doped Nanocrystalline Diamond Films as Reflective Layers for Fiber-Optic Sensors of Refractive Index of Liquids

Monika Kosowska ¹, Daria Majchrowicz ¹, Kamatchi J. Sankaran ^{2,3}, Mateusz Ficek ¹, Ken Haenen ^{2,3} and Małgorzata Szczerska ^{1,*}

¹ Department of Metrology and Optoelectronics, Faculty of Electronics, Telecommunications and Informatics, Gdansk University of Technology, 11/12 Narutowicza Street, 80-233 Gdansk, Poland

² Institute for Materials Research (IMO), Hasselt University, Wetenschapspark 1, B-3590 Diepenbeek, Belgium

³ IMOMEC, IMEC vzw, Wetenschapspark 1, B-3590 Diepenbeek, Belgium

* Correspondence: mjedrzej@eti.pg.edu.pl; Tel.: +48-583-471-361

Received: 28 May 2019; Accepted: 29 June 2019; Published: 2 July 2019



Abstract: This paper reports the application of doped nanocrystalline diamond (NCD) films—nitrogen-doped NCD and boron-doped NCD—as reflective surfaces in an interferometric sensor of refractive index dedicated to the measurements of liquids. The sensor is constructed as a Fabry–Pérot interferometer, working in the reflective mode. The diamond films were deposited on silicon substrates by a microwave plasma enhanced chemical vapor deposition system. The measurements of refractive indices of liquids were carried out in the range of 1.3 to 1.6. The results of initial investigations show that doped NCD films can be successfully used in fiber-optic sensors of refractive index providing linear work characteristics. Their application can prolong the lifespan of the measurement head and open the way to measure biomedical samples and aggressive chemicals.

Keywords: doped nanocrystalline diamond films; refractive index sensor; fiber-optic; nitrogen-doping; boron-doping; optical fiber sensor

1. Introduction

The field of biomedical measurements is rapidly growing, and optical sensors play a significant role in its development. The use of optical fibers in the construction of sensors provides many advantages [1]. Such sensors require only a small amount of sample, give a rapid response, measurements are non-invasive and chemical pretreatment is not needed [2]. Moreover, they assure no risk of electrical sparks and immunity to ionizing radiation [3]. This group of sensors is widely used for the determination of refractive index, one of the most important optical properties describing materials. The precise determination of its value allows for the identification of the investigated substance and its concentration [4,5]. Therefore, fiber-optic measurements of refractive index attracted considerable attention in finding applications in many fields including biomedicine, chemistry, environmental analysis and the food industry [6–10].

However, measurements of biological samples, hazardous and chemically aggressive substances are still a challenge. Investigation of such materials carries a high risk of damaging the elements of the measurement heads. In conventional Fabry–Pérot interferometers, the most exposed part is a mirror that has direct contact with the sample. In standard solutions it is made of metallic layers such as silver or aluminum. Even though they have satisfactory optical parameters, they are susceptible to mechanical damage and have limited chemical resistance. Aluminum can cause poisoning of the investigated tissues. This results in permanent damaging of the measurement head and the necessity

of replacing the damaged part, which is inconvenient and causes additional costs. In performing measurements of biomedical substances, apart from the risk of damaging the sensor, there is a high chance of polluting the investigated sample, which can affect the results giving a false response. Thus, there is a need for new materials which can be used in the production of reflective surfaces that provide good optical parameters, as well as great resistance to chemicals. These features can be found in diamond materials.

Diamond films deposited in chemical vapor deposition (CVD) process are very hard, chemically inert and stable materials, which are transparent in a broad wavelength range [11–13]. The properties of diamond film structures produced in a CVD system can be tailored by changing the deposition process parameters. The use of dopants in the working gas mixture has an impact on the resulting film, including the optical properties. Hence it is possible to tune the process to achieve a suitable structure for the reflective layer in the interferometer [14–16].

In this work we present the application of diamond films doped with boron and nitrogen. The diamond films were deposited on silicon substrates and utilized as mirrors in fiber-optic sensors dedicated to measuring refractive indices of liquids. This solution combines the advantages of the fiber-optic Fabry–Pérot interferometer with benefits introduced by the extraordinary parameters of the diamond films, solving the problem of standard mirrors being susceptible to chemical damage.

This study is a continuation of our investigation regarding the use of diamond films in fiber-optic sensors. Previously, we proved that doped diamond films can be successfully applied in distance sensors and that they assure the biocompatibility [17,18]. This paper reports refractive index sensors utilizing two kinds of doped diamond films.

2. Materials and Methods

2.1. Nanocrystalline Diamond Films

Boron-doped nanocrystalline diamond (BD-NCD) films were synthesized in a microwave plasma enhanced CVD (MW PE CVD) system (SEKI Technotron AX5400S, Tokyo, Japan) on p-type Si (100) substrates. In our experiments, dedicated NCD suspensions were used to seed the substrates [19,20]. The pressure before growth in the vacuum chamber was kept at 10^{-4} Torr. A special truncated cone-shaped shim and 500 °C was used during the growth of BD-NCD films. The plasma microwave power, optimized for diamond synthesis, was kept at 1300 W [21]. In this study, the molar ratio of the CH₄:H₂ mixture was kept at 1% of gas volume at 300 sccm of the total flow rate. The boron level, expressed as the (B)/(C) ratio, in the gas phase was 10,000 ppm. The growth time was 3 h. After the growth process, the substrate temperature was slowly reduced ($5\text{ °C}\cdot\text{min}^{-1}$) down to room temperature.

Nitrogen-doped NCD (ND-NCD) film was grown on silicon substrate in a MW PE CVD system (SEKI ASTeX 6500, Tokyo, Japan). Prior to the diamond film growth, the silicon substrate was seeded with a colloidal suspension containing 5 nm detonation nanodiamonds and distilled water using a spin-coating technique. The ND-NCD film was grown on a silicon substrate in a CH₄(9)/H₂ (282)/N₂(3) sccm plasma excited by 3000 W microwave power with 65 Torr pressure for 180 min. The substrate temperature was about 650 °C, which was measured using a single-color optical pyrometer.

The samples were made in two different universities. Two different pressures result from differences in the standard growth procedures developed in each scientific team. For convenient referencing to each sample, we adopted the following abbreviations: BD-NCD-Si—boron-doped nanocrystalline diamond deposited on a silicon substrate, and ND-NCD-Si—nitrogen-doped nanocrystalline diamond deposited on a silicon substrate.

Diamond Film Characterization

In order to be able to use a diamond film as a mirror, it has to fulfill specific requirements considering its surface. The film has to be homogenous, continuously covering the substrate. No cracks should be detected to provide reliable resistance to chemicals.

The diamond films used in this experiment were previously applied in fiber-optic sensors for determining their potential in measurements of biological samples and for distance measurements [14,22], hence, their detailed characterizations can be found there. The investigation has shown that the thickness of the doped NCD films is around 300 nm, and the root mean square roughness is equal to 22 nm and 49 nm for BD-NCD-Si and ND-NCD-Si, respectively [22]. The aforementioned examinations proved that the deposited diamond films can work as reflective surfaces.

2.2. Measurement Setup

The fiber-optic setup for measuring the refractive index of liquid samples based on a Fabry–Pérot interferometer was constructed. The experiment was carried out with two superluminescent diodes working at the central wavelength of 1290 nm and 1560 nm, respectively. The system was built with a light source (S-1290-G-I-20, and S1550-G-I-10, Superlum, Ireland), an optical spectrum analyzer (ANDO AQ6319, Yokogawa, Japan), single-mode telecommunication optical fibers (SMF-28, Thorlabs, Newton, NJ, USA), and a 2×1 50:50 coupler. The tip of the fiber and the doped NCD films were used as a measurement head. The principle of the operation of this setup was described elsewhere [22]. Figure 1 shows the schema of the measurement setup with a close-up of the measurement head.




Figure 1. Measurement setup with a close-up of the measurement head.

The measurement field, called a Fabry–Pérot cavity, is formed between the face of a fiber and the nanocrystalline diamond film deposited on silicon substrate. Measurements of refractive indices in the range of 1.3–1.6 were performed with the use of refractive index liquids provided by Cargille®. In the experiment and data analysis we considered the refractive indices of liquids at 1290 nm and 1560 nm according to the datasheets. However, in the text, we refer to each solution by a rounded value of its refractive index to provide greater clarity. Before placing the sample, the cavity was cleaned with isopropyl alcohol. The block diagram of the experiment is shown in Figure 2.




Figure 2. The block diagram of the experiment.

The distance between the fiber and the reflective surface in the device influences the visibility of the interferometric fringes, as the light beam diverges and is coupled back into the fiber with differing efficiency. Therefore, before carrying out refractive index measurements, the cavity length of the

interferometer was set to achieve the highest visibility value of the measured signal. To calculate the visibility V , the following formula was used:

$$V = \frac{I_{\max} - I_{\min}}{I_{\max} + I_{\min}} \quad (1)$$

where I_{\max} is the maximum intensity of the measured signal and I_{\min} is the minimum intensity of the measured signal. Table 1 shows values of visibility for the investigated films calculated for representative cavity lengths for given working wavelengths.

Table 1. Visibility values calculated for representative cavity lengths filled with air measured on BD-NCD-Si and ND-NCD-Si for central wavelength of 1290 nm and 1560 nm.

BD-NCD-Si			
Wavelength—1290 nm		Wavelength—1560 nm	
Cavity Length	Visibility	Cavity Length	Visibility
80 μm	0.9915	110 μm	0.9950
160 μm	0.8239	200 μm	0.8727
ND-NCD-Si			
Wavelength—1290 nm		Wavelength—1560 nm	
Cavity Length	Visibility	Cavity Length	Visibility
60 μm	0.9917	90 μm	0.9939
120 μm	0.8394	190 μm	0.7905

BD-NCD-Si: boron-doped nanocrystalline diamond deposited on a silicon substrate; ND-NCD-Si: nitrogen-doped nanocrystalline diamond deposited on a silicon substrate.

The above calculations of visibility were based on spectra presented in Figures 3 and 4.




Figure 3. Spectra measured on BD-NCD-Si while the cavity was filled with air. Spectra obtained for 1290 nm and cavity lengths (a) 80 μm and (b) 160 μm . Spectra obtained for 1560 nm and cavity lengths (c) 110 μm and (d) 200 μm .




Figure 4. Spectra measured on ND-NCD-Si while the cavity was filled with air. Spectra obtained for 1290 nm and cavity lengths (a) 60 μm and (b) 120 μm . Spectra obtained for 1560 nm and cavity lengths (c) 90 μm and (d) 190 μm .

3. Results

The measurement signals were recorded with the use of an optical spectrum analyzer. The representative spectra are shown in Figure 5.




Figure 5. Spectra of liquid with refractive index $n = 1.4$ measured with a wavelength of 1290 nm (a) BD-NCD-Si, (b) ND-NCD-Si.

The obtained spectra were analyzed regarding the spectral separation between the maxima in the spectra for measured refractive indices of liquids. The results of these investigations for the light sources with a central wavelength equal to 1290 nm are presented in Figure 6.






Figure 6. Measurement results: spectral separation between maxima as a function of refractive index for wavelength equal to 1290 nm (a) BD-NCD-Si, (b) ND-NCD-Si.

The values of correlation coefficient R^2 and sensitivity S for each sensor working at central wavelength of 1290 nm are presented in Table 2. The sensitivity can be directly found from the slope of the linear fit [9]. The values of R^2 parameter in the range of 0.7–0.9 indicate high positive/negative correlation, while values higher than 0.9 mean very high positive/negative correlation [23].

Table 2. Correlation coefficient (R^2) and sensitivity (S) values for sensors working at 1290 nm.

Parameters	BD-NCD-Si	ND-NCD-Si
R^2	0.8950	0.7064
S (nm/a.u.)	−4.6658	−6.0187

Similarly, the same procedure of data analysis was applied to the measurement data collected with the use of a source working at central wavelength of 1560 nm. The representative measurement spectra are shown in Figure 7.




Figure 7. Spectra of liquid with refractive index $n = 1.6$ measured with a wavelength of 1560 nm (a) BD-NCD-Si, (b) ND-NCD-Si.

Figure 8 presents the relationship between the spectral separation of the interferometric fringes and the refractive index.




Figure 8. Measurement results: spectral separation between maxima as a function of refractive index for wavelength equal to 1560 nm (a) BD-NCD-Si, (b) ND-NCD-Si.

The values of correlation coefficient R^2 and sensitivity S for each sensor working at central wavelength of 1560 nm are calculated and presented in Table 3.

Table 3. Correlation coefficient (R^2) and sensitivity (S) values for sensors working at 1560 nm.

Parameters	BD-NCD-Si	ND-NCD-Si
R^2	0.9513	0.8494
S (nm/a.u.)	-7.2031	-5.8673

It can be noted that all investigated samples can be applied as reflective surfaces in fiber-optic sensors of refractive index. The dependence between the spectral separation and refractive index is linear in each case, with a high or very high value of correlation coefficient R^2 .

The small reduction in sensitivity for the nitrogen-doped NCD can result from the lower value of reflectivity assured by this film in comparison with the boron-doped NCD film. The sensitivity is directly related to the reflection assured by the reflective film implemented in the sensor. The reflectivity of the surface is dependent on its refractive index. Reflectivity R at the boundary between the diamond film and the medium can be calculated by use of the Fresnel equation [24]:

$$R = \left(\frac{n_2 - n_1}{n_1 + n_2} \right)^2 \quad (2)$$

where: n_1 , n_2 are the refractive indices of the medium and NCD diamond film, respectively.

For the BD-NCD-Si sample, the value of the refractive index increases with the increase of wavelength, which results in a higher value of reflectivity. For the ND-NCD-Si sample, the value of the refractive index decreases with the increase of wavelength, which results in a lower value of reflectivity. The reflectivity for the BD-NCD-Si sample increased by 3.6% with the wavelength, while the reflectivity of the ND-NCD-Si sample decreased by 0.79%.

4. Conclusions

In this paper we reported the application of doped NCD films to fiber-optic interferometric sensors. The sensors are dedicated to the measurement of the refractive index of liquids. The experiments included NCD films doped with nitrogen and boron. The diamond films were produced using a microwave plasma enhanced chemical vapor deposition system. The investigation was performed in the refractive index range between 1.3–1.6. The linear mathematical models were adopted to the measurement data, allowing for determination of correlation coefficients R^2 and sensitivity values for

each sensor. The achieved values of R^2 in the range of 0.7–0.9 and higher than 0.9 show high and very high negative correlation between the spectral separation of the fringes in the spectrum and the value of the refractive index. The initial study showed that the use of doped diamond films as reflective surfaces in sensors of refractive index is possible. This approach assures better resistance to chemical and mechanical damage prolonging the lifespan of sensors and allowing measurements of aggressive chemicals and biological samples.

Author Contributions: Conceptualization M.K., D.M. and M.S.; methodology D.M. and M.S.; validation M.K.; data analysis M.K. and D.M.; investigation M.K.; samples preparation K.J.S., M.F. and K.H.; data curation M.K.; writing—original draft preparation M.K., K.J.S., D.M., M.F., K.H. and M.S.; writing—review and editing K.H. and M.S.; visualization M.K.; supervision, K.H. and M.S.; project administration M.S. and D.M.; funding acquisition M.K., D.M. and M.S.

Funding: The authors want to acknowledge the financial support of the Polish National Science Centre under Grant No. 2017/25/N/ST7/01610, the Polish National Centre for Research and Development under the project Techmatstrateg Diamsec 347324, DS Programs of Faculty of Electronics, Telecommunications and Informatics of Gdańsk University of Technology and the project no MNiD/2018/2019/4. M.S. acknowledges the scholarship from the Bekker Programme PPN/BEK/2018/1/00185 funded by the Polish National Agency for Academic Exchange. K.J.S. and K.H. want to acknowledge the financial support from the Methusalem “NANO” network.

Conflicts of Interest: The authors declare no conflict of interest.

References

- Chiavaioli, F.; Baldini, F.; Tombelli, S.; Trono, C.; Giannetti, A. Biosensing with optical fiber gratings. *Nanophotonics* **2017**, *6*, 663–679. [[CrossRef](#)]
- Chaudhari, A.L.; Shaligram, A.D. Fiber optic sensor for the measurement of concentration and refractive index of liquids based on intensity modulation. *Int. J. Mod. Phys. Conf. Ser.* **2012**, *6*, 589–593. [[CrossRef](#)]
- Ahsani, V.; Ahmed, F.; Jun, M.B.G.; Bradley, C. Tapered Fiber-Optic Mach-Zehnder Interferometer for Ultra-High Sensitivity Measurement of Refractive Index. *Sensors* **2019**, *19*, 1652. [[CrossRef](#)] [[PubMed](#)]
- Zhang, T.; Feng, G.; Song, Z.; Zhou, S. A single-element interferometer for measuring refractive index of transparent liquids. *Opt. Commun.* **2014**, *332*, 14–17. [[CrossRef](#)]
- Kaneko, K.; Yoshimura, Y.; Shimizu, A. Water concentration dependence of the refractive index of various ionic liquid-water mixtures. *J. Mol. Liquids* **2018**, *250*, 283–286. [[CrossRef](#)]
- Rodionov, S.A.; Remnev, M.A.; Klimov, V.V. Refractive index sensor based on all-dielectric gradient metasurface. *Sens. Bio-Sens. Res.* **2019**, *22*, 100263. [[CrossRef](#)]
- Nath, P.; Singh, H.K.; Datta, P.; Sarma, K.C. All-fiber optic sensor for measurement of liquid refractive index. *Sens. Actuators A Phys.* **2008**, *148*, 16–18. [[CrossRef](#)]
- Yüksel, K. Optical fiber sensor system for remote and multi-point refractive index measurement. *Sens. Actuators A Phys.* **2016**, *250*, 29–34. [[CrossRef](#)]
- Chiavaioli, F.; Gouveia, C.A.J.; Jorge, P.A.S.; Baldini, F. Towards a Uniform Metrological Assessment of Grating-Based Optical Fiber Sensors: From Refractometers to Biosensors. *Biosensors* **2017**, *7*, 23. [[CrossRef](#)]
- Chiavaioli, F.; Zubiarte, P.; Del Villar, I.; Zamarreño, C.R.; Giannetti, A.; Tombelli, S.; Trono, C.; Arregui, F.J.; Matias, I.R.; Baldini, F. Femtomolar Detection by Nanocoated Fiber Label-Free Biosensors. *ACS Sens.* **2018**, *3*, 936–943. [[CrossRef](#)]
- Sobaszek, M.; Skowroński, Ł.; Bogdanowicz, R.; Siuzdak, K.; Cirocka, A.; Zięba, P.; Gnyba, M.; Naparty, M.; Gołuński, Ł.; Płotka, P. Optical and electrical properties of ultrathin transparent nanocrystalline boron-doped diamond electrodes. *Opt. Mater.* **2015**, *42*, 24–34. [[CrossRef](#)]
- Ficek, M.; Sobaszek, M.; Gnyba, M.; Ryl, J.; Gołuński, Ł.; Smietana, M.; Jasiński, J.; Caban, P.; Bogdanowicz, R. Optical and electrical properties of boron doped diamond thin conductive films deposited on fused silica glass substrates. *Appl. Surf. Sci.* **2016**, *387*, 846–856. [[CrossRef](#)]
- Ficek, M.; Sankaran, K.J.; Ryl, J.; Bogdanowicz, R.; Lin, I.-N.; Haenen, K.; Darowicki, K. Ellipsometric investigation of nitrogen doped diamond thin films grown in microwave CH₄/H₂/N₂ plasma enhanced chemical vapor deposition. *Appl. Phys. Lett.* **2016**, *108*, 241906. [[CrossRef](#)]
- Majchrowicz, D.; Kosowska, M.; Struk, P.; Jędrzejewska-Szczerska, M. Tailoring the Optical Parameters of Optical Fiber Interferometer with Dedicated Boron-Doped Nanocrystalline Diamond Thin Film. *Phys. Status Solidi A* **2017**, *214*, 1700222. [[CrossRef](#)]

15. Milewska, D.; Karpienko, K.; Jędrzejewska-Szczerska, M. Application of thin diamond films in low-coherence fiber-optic Fabry P erot displacement sensor. *Diam. Mater.* **2016**, *64*, 169–176. [[CrossRef](#)]
16. Kosowska, M.; Majchrowicz, D.; Sankaran, K.J.; Ficek, M.; Jędrzejewska-Szczerska, M.; Haenen, M.K. Nitrogen-doped diamond thin films: Potential application in Fabry-P erot interferometer. *Proc. SPIE* **2017**, *10716*, 1071614.
17. Majchrowicz, D.; Kosowska, M.; Sankaran, K.J.; Struk, P.; W asowicz, M.; Sobaszek, M.; Haenen, K.; Jędrzejewska-Szczerska, M. Nitrogen-Doped Diamond Film for Optical Investigation of Hemoglobin Concentration. *Materials* **2018**, *11*, 109. [[CrossRef](#)]
18. Jędrzejewska-Szczerska, M.; Majchrowicz, D.; Hirsch, M.; Struk, P.; Bogdanowicz, R.; Bechelany, M.; Tuchin, V.V. *Nanotechnology and Biosensors: Nanolayers in Fiber-Optic Biosensing*; Elsevier: Amsterdam, The Netherlands, 2018.
19. Ficek, M.; Bogdanowicz, R.; Ryl, J. Nanocrystalline CVD Diamond Coatings on Fused Silica Optical Fibers: Optical Properties Study. *Acta Phys. Polonica A* **2015**, *127*, 868–873. [[CrossRef](#)]
20. Bogdanowicz, R.;  mietana, M.; Gnyba, M.; Ficek, M.; Stra n ak, V.; Golu nski,  .; Sobaszek, M.; Ryl, J. Nucleation and growth of CVD diamond on fused silica optical fibres with titanium dioxide interlayer. *Phys. Status Solidi A* **2013**, *210*, 1991–1997. [[CrossRef](#)]
21. Dec, B.; Ficek, M.; Ryciewicz, M.; Macewicz,  .; Gnyba, M.; Sawczak, M.; Sobaszek, M.; Bogdanowicz, R. Gas Composition Influence on the Properties of Boron-Doped Diamond Films Deposited on the Fused Silica. *Mater. Sci.-Poland* **2018**, *36*, 288–296. [[CrossRef](#)]
22. Hirsch, M.; Majchrowicz, D.; Wierzba, P.; Weber, M.; Bechelany, M.; Jędrzejewska-Szczerska, M. Low-Coherence Interferometric Fiber-Optic Sensors with Potential Applications as Biosensors. *Sensors* **2017**, *17*, 261. [[CrossRef](#)] [[PubMed](#)]
23. Mukaka, M. A guide to appropriate use of Correlation coefficient in medical research. *Malawi Med J.* **2012**, *24*, 69–71. [[PubMed](#)]
24. Grattan, L.S.; Meggitt, B.T. *Optical Fiber Sensor Technology*; Springer: Boston, MA, USA, 2000.



  2019 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).

