



Review

Long-Period Gratings and Microcavity In-Line Mach Zehnder Interferometers as Highly Sensitive Optical Fiber Platforms for Bacteria Sensing

Tinko Eftimov ^{1,*},†, Monika Janik ^{2,3,†} , Marcin Koba ⁴ , Mateusz Śmietana ², Predrag Mikulic ¹ and Wojtek Bock ¹

¹ Centre de Recherche en Photonique, Université du Québec en Outaouais, 101 rue Saint-Jean-Bosco, Gatineau, QC J8X 3X7, Canada; predrag.mikulic@uqo.ca (P.M.); wojtek.bock@uqo.ca (W.B.)

² Institute of Microelectronics and Optoelectronics, Warsaw University of Technology, Koszykowa 75, 00-662 Warsaw, Poland; monika.janik@pw.edu.pl or monika.janik@pg.edu.pl (M.J.); m.smietana@elka.pw.edu.pl (M.Ś.)

³ Department of Metrology and Optoelectronics, Faculty of Electronics, Telecommunications and Informatics, Gdansk University of Technology, Narutowicza 11/12, 80-233 Gdansk, Poland

⁴ National Institute of Telecommunications, Szachowa 1, 04-894 Warsaw, Poland; mkoba@elka.pw.edu.pl

* Correspondence: tinko.eftimov@uqo.ca

† Both authors contributed equally to this work.

Received: 18 May 2020; Accepted: 27 June 2020; Published: 5 July 2020



Abstract: Selected optical fiber sensors offer extraordinary sensitivity to changes in external refractive (RI), which make them promising for label-free biosensing. In this work the most sensitive ones, namely long-period gratings working at (DTP-LPG) and micro-cavity in-line Mach-Zehnder interferometers (μ IMZI) are discussed for application in bacteria sensing. We describe their working principles and RI sensitivity when operating in water environments, which is as high as 20,000 nm/RIU (Refractive index unit) for DTP-LPGs and 27,000 nm/RIU for μ IMZIs. Special attention is paid to the methods to enhance the sensitivity by etching and nano-coatings. While the DTP-LPGs offer a greater interaction length and sensitivity to changes taking place at their surface, the μ IMZIs are best suited for investigations of sub-nanoliter and picoliter volumes. The capabilities of both the platforms for bacteria sensing are presented and compared for strains of *Escherichia coli*, lipopolysaccharide *E. coli*, outer membrane proteins of *E. coli*, and *Staphylococcus aureus*. While DTP-LPGs have been more explored for bacteria detection in 10^2 – 10^6 Colony Forming Unit (CFU)/mL for *S. aureus* and 10^3 – 10^9 CFU/mL for *E. coli*, the μ IMZIs reached 10^2 – 10^8 CFU/mL for *E. coli* and have a potential for becoming picoliter bacteria sensors.

Keywords: optical fiber sensors; label-free biosensing; bacteria detection; long-period gratings; microcavity Mach-Zehnder interferometers

1. Introduction

Bacteria have been a crucial target for rapid detection and identification in fields, such as food and water safety, medicine, or public health and security. Bacterial infectious diseases are one of the leading causes of deaths and hospitalizations worldwide [1]. For medical diagnostics and targeted treatment, a rapid identification of the bacterial strain is crucial. Moreover, during bacterial outbreaks, such examination can provide key information that can help to decrease the spread of the infection. Nowadays, conventional methods for bacteria-specific detection relay mostly on culture methods in selective media [2], immunological tests such as enzyme-linked immunosorbent assays (ELISAs) [3], or molecular techniques, such as real-time PCR [4]. Even though they are very effective, all of them

are time-consuming, expensive, require extensive sample preparation, specialized equipment and technical staff to proceed with the tests.

Thus, there is an urgent need for more rapid, cost-effective, easy-to-use, sensitive tests, which could efficiently identify bacteria outside of the laboratory, or at the point of care. The future biosensors should be compact, operate at minimal sample volume, inexpensive, easy to work with, highly specific and sensitive. Moreover, they should be able to operate based on label-free methods, as these are relatively fast and do not need additional labeling/preparation of the tested sample.

Most of the optical fiber biosensors exploit the label-free target binding-induced changes, which alternate the sensor surface's refractive index (RI), and the information is then transduced to a detector and is typically identified as alteration of the output spectrum. For the purposes of biosensing, the changes in correspond to binding of a biological target, e.g., biomolecules or microorganisms, such as bacteria. Out of a great variety of optical fiber RI sensors these based on long-period gratings (LPGs) [5–7] and different versions of Mach-Zehnder interferometers [8] can be pointed out as the most sensitive.

The RI sensitivity of the LPG depends on RI range, and for commonly investigated LPGs is high around RI of fiber cladding, and fairly low around RI of water, where it is expected to be high when bacteria sensing is considered. For these LPGs, thin, high RI overlays may largely increase the sensitivity up to about 2000 nm/RIU [9] (RIU—Refractive Index Units). LPGs working in proximity of dispersion turning point (DTP) where double resonance (DR) effect can be observed, have a total sensitivity of both minima up to 800 nm/RIU around water and up to 10,800 nm/RIU in the range of 1.4436–1.4489 [10]. DTP-LPGs offer two resonance minima shifting towards opposite directions with RI increase. The DTP working conditions can be fine-tuned by etching the cladding [11]. DTP-LPGs are the most sensitive to RI changes, which makes them suitable for biosensing applications [12,13].

There is a wide selection of optical fiber Mach-Zehnder interferometers, which vary by fabrication technology, sensing length, type of interference, and sensitivity to RI [14]. A microcavity drilled across the fiber represents the most compact structure [15] referred to as micro in-line Mach-Zehnder Interferometer (μ IMZIs). A Mach-Zehnder interferometer based on core cladding intermodal interference in photonic crystal fibers (PCF) has shown a sensitivity of 100–150 nm/RIU [16]. Other types of RI sensors are those based on mode coupling along with a two-core fiber exhibiting a sensitivity as high as 2100 nm/RIU [17], while RI sensors with asymmetrical tapered fiber based on evanescent field reaches as high as 3390–3914 nm/RIU [18]. Another approach is an interference in a single mode [19] and loss change in multimode [20] D-shaped fibers. In nanocoated D-shaped single mode sensitivity can reach up to 1566 nm/RIU in the 1.3–1.335 range [19]. Another structure is a micro-spiral upon a single-multi-single mode fiber section for which the RI sensitivity of 2114 nm/RIU in the 1.3373–1.4345 range can be achieved [21].

In the present paper, we analyze the capabilities of two most surrounding refractive index (SRI) sensitive structures namely DR-LPGs and μ IMZIs, showing their applicability for bacteria sensing.

2. DTP-LPGs as Highly RI Sensitive Optical Fiber Platform

As outlined in the introduction LPGs, especially DTP-LPGs, exhibit high RI sensitivities and as such represent a desirable platform for bacteria and label-free biosensing in general. LPG is a periodic modulation of the RI introduced along a core of a single-mode optical fiber. The modulation causes a part of the power of the fundamental mode linearly polarized LP_{01} to be coupled to a higher order cladding mode LP_{0p} and they exchange power along the structure (Figure 1a). The propagation constants of the modes are defined as $\beta_1 = n_1 2\pi/\lambda$ and $\beta_p = n_p 2\pi/\lambda$ for fundamental and high order modes, respectively, where n_1 and n_p are effective RI of core and cladding, respectively. Coupling from the core to the cladding modes occurs when the phase matching condition $\delta\beta = 2\pi/\Lambda$ is met, where Λ is a period the modulation [22]. The resonance wavelength λ_c at which the phase matching condition is fulfilled is found as in Equation (1).

$$\lambda_c = (n_1 - n_p)\Lambda = \Delta n_{eff}\Lambda \quad (1)$$

Since higher order modes are strongly attenuated in the cladding, only the power remaining in the core mode is measured. At resonance conditions most of the power is coupled to the lossy higher order mode LP_{0p} so a minimum in the transmission is observed at λ_c (Figure 1a). The spectral position of this minimum depends on changes of surrounding RI, as well as temperature, axial strain, bending, and twisting. Bio-sensing applications of LPGs make use of the RI sensitivity which is very high around RI of fiber cladding ($n = 1.445$), but it is relatively low around RI of water ($n = 1.333$). Since bacteria stay alive in a water medium, the RI sensitivity of LPGs in proximity to that of water is expected. LPGs around DTP offer this property.




Figure 1. A long-period gratings (LPG) and its responses: (a) coupling between the fundamental LP_{01} core mode and a higher order LP_{0p} cladding mode along the grating and typical transmission spectrum, away from dispersion turning point (DTP) with a minimum at resonance wavelength λ_c . (b) The relation between the period Λ and the resonance wavelength λ_0 of an LPG around the DTP at a period Λ_0 (lower curve). Changing the period Λ of the grating to Λ_A and Λ_B , the spectrum of the LPGs splits and shifts in opposite directions as shown in (b) above.

2.1. LPGs around the DTP

As seen in Equation (1) the resonance wavelength is proportional to the grating period. However, the effective RI difference Δn_{eff} is not constant with respect to wavelength λ . Because of Δn_{eff} dispersion, the grating period and the resonance wavelength for a given higher order cladding mode are related in a way as shown in the lower part of Figure 1b. For a given LP_{0p} mode at a period Λ_0 there is only one minimum. This marks the so called dispersion turning point (DTP). For smaller values of the period ($\Lambda < \Lambda_0$) there are two resonance wavelengths: a lower λ_L and a higher λ_H ($\lambda_{A,L} < \lambda_{A,H}$), which means that there are two different values of Δn_{eff} for each resonance wavelength, namely $\Delta n_{L} < \Delta n_{H}$. As seen from Figure 2b, as the period Λ decreases, λ_L decreases while λ_H increases. On the other side, once the grating is fabricated Λ is constant and if for some reason Δn_{eff} changes, then so do the two resonance wavelengths accordingly. This occurs if for example strain, temperature, or surrounding RI changes are introduced upon the grating.

Since DTP-LPGs are the most sensitive to RI, they are fabricated in a way as to be just before DTP when surrounded by air and to appear in water with a minimum at λ_0 . In such a manner, any change of RI above that of water will be detected as an increase of losses or a split in the spectrum with minima at wavelengths λ_L and λ_H going apart. The LPG thus becomes a double-resonance grating. The total wavelength shift between the two resonances will be then defined as in Equation (2).

$$\Delta\lambda = \lambda_H - \lambda_L \quad (2)$$

In this paper, we make use of this effect to obtain maximum sensitivity of the LPG around DTP. The responses to RI of two DTP-LPGs are presented below. The first with a period of $\Lambda = 135.6 \mu\text{m}$ at a center wavelength $\lambda_c = 770 \text{ nm}$ was UV written in a Fibercore SM600 single mode fiber with a cut-off wavelength in the 500–600 nm range. The second with a period of $\Lambda = 222.8 \mu\text{m}$ at a center wavelength $\lambda_c = 1583.4 \text{ nm}$ was UV written in a Corning SMF-28 (cut-off at 1260 nm). The UV laser was a Krypton fluoride Lumonics KrF excimer laser. Both gratings fabricated at the labs of the Centre de Recherche en Photonique-UQO were etched using hydrofluoric HF acid to fine-tune them around the turning point in a way as to observe it in water immersion.




Figure 2. Spectral response of DTP-LPGs to refractive (RI) changes: (a) SM600-based LPG in the visible/near infrared (Vis DTP-LPG); (b) SMF-28-based LPG in the infrared (IR) DTP-LPG.

For the DTP-LPGs, two types of spectral changes are observed. The first (I) are the resonance wavelength shifts in opposite directions of λ_L and λ_H and hence the increase of $\Delta\lambda$ with RI. The second (II) is the increase of the power level (or loss decrease) ΔI_{DTP} at λ_{DTP} . In both the $\Delta\lambda$ vs. RI and the ΔI_{DTP} vs. RI dependences, we note two ranges: the first one is for RI from 1.33 to 1.34 offering higher sensitivity when compared to the second range for RI > 1.34 exhibiting a lower sensitivity.

The Vis DTP-LPG around RI of water RI exhibits sensitivity of about 11,000 nm/RIU which falls to about 690 nm/RIU away from it (Figure 3a), while the IR DTP-LPG offers a sensitivity of about 6300 nm/RIU around RI of water and about 2700 nm/RIU away from it (Figure 4a). Similarly, in terms of change in power at fixed wavelength, the Vis DTP-LPG shows higher RI sensitivity of about 970 dB/RIU for RI below 1.34 that is reduced to 124 dB/RIU beyond that (Figure 3b).




Figure 3. Responses of the Vis DTP-LPG: (a) $\Delta\lambda$ vs. RI; (b) ΔI_{DTP} vs. RI.




Figure 4. Responses of the IR DTP-LPG: (a) $\Delta\lambda$ vs. RI; (b) ΔI_{DTP} vs. RI.

The IR DTP-LPG has the same sensitivity in the whole RI range reaching about 317 dB/RIU (Figure 4b). In both the DTP cases the sensitivities around RI of water is much high as it is expected for biosensing applications.

2.2. Fabrication of LPGs

2.2.1. Fabrication Technologies

LPGs can be arc-induced using a fiber splicer [23], fabricated using UV pulsed [24], femtosecond laser or CO₂ laser. Comparative analysis shows that UV-written and arc-induced LPGs may offer similar sensitivities [25]. Out of the mentioned above methods, the UV laser method uses amplitude masks while the others are point-by-point writing methods. The arc-induced method is the simplest and the LPGs spectrum is stable in time and temperature. However, the minimum periods of the gratings are limited by the electric arc dimensions to about 180 μm . Moreover, the repeatability of their parameters is not as good as of the LPGs written with amplitude masks and UV lasers. Amplitude mask and point-by-point laser written LPGs are not limited in period length, but need additional post-fabrication annealing to stabilize the grating parameters in the long term. All point-by-point writing methods allow greater flexibility in medication of the grating periodicity or coupling strength along the structure. The described biosensors were fabricated using KrF excimer laser in combination with an amplitude mask and a subsequent etching.

2.2.2. Fine-Tuning Approaches

The fine-tuning of the LPGs includes, basically, the tailoring of the sensitivity and the tuning of the LPG to a definite initial spectral distribution close to the DTP as shown above. This is basically done by cladding etching and nanocoating with, e.g., Al₂O₃ [26,27]. Sensitivities as high as 20,000 nm/RIU around RI of water have been achieved and reported [26]. The deposition of Al₂O₃ has been found also to cause center wavelength shifts [27]. Due to this, the sensitivity and the center wavelength of the DTP-LPG can be fine-tuned. However, since DTP-LPGs are highly sensitive not only to RI, but also to strain, temperature twists and bends, experimental difficulties with noise reduction arises in measurements with these gratings, which can be overcome by isolating the sensing grating from unwanted stimuli [28].

2.2.3. Temperature Stability

Since DTP-LPGs are strongly temperature dependent, to ensure temperature stability LPGs are placed on temperature stabilized conditions or in glass tubes to eliminate temperature fluctuations caused by air flows. A more sophisticated method is based on a cascaded structure of two DTP-LPGs

separated by a fiber section, which ensures a temperature stability of $-0.45 \text{ pm}/^\circ\text{C}$ and an RI sensitivity of $2583.3 \text{ nm}/\text{RIU}$ in the $1.333\text{--}1.343$ range [29]. The structure is in fact a Mach-Zehnder interferometer and the intermediate section introduces a temperature dependent phase shift compensating the LPG temperature dependence.

3. Micro In-Line Mach-Zehnder Interferometers as Highly RI Sensitive Optical Fiber Platforms

3.1. In-Line All Fiber Mach-Zehnder Interferometers

As outlined in the introduction, the other highly RI sensitive and thus promising fiber structure for biosensing, including bacteria sensing, are all fiber-based Mach-Zehnder interferometers (MZIs). Comparing to LPGs where the sensor length is up to ca. 50 mm, the lengths of femtosecond laser fabricated micro-cavity MZIs can be as short as $10 \text{ }\mu\text{m}$ [14,30]. Tapered MZIs have sensing lengths from 30 mm to 40 mm and are comparable to the lengths of LPGs. Their sensitivities to RI changes are of the order of $400 \text{ nm}/\text{RIU}$ around RI of water [31], which is considerably less than that of DTP-LPGs. A MZI based on up-down tapers [32] offers a sensitivity of $-310 \text{ dB}/\text{RIU}$, which again is comparable or less than a similar type of sensitivity of DTP-LPGs where it varies from $317 \text{ dB}/\text{RIU}$ to $970 \text{ dB}/\text{RIU}$ (Figures 3b and 4b). A somewhat better sensitivity of $703 \text{ nm}/\text{RIU}$ is reported for a multicore tapered sensor of 20 mm [33]. MZIs based on etched photonic crystal fiber sandwiched between standard SMF fibers exhibited a sensitivity of $360 \text{ nm}/\text{RIU}$ for a 35 mm long section length and PCF fiber cladding reduced to $91 \text{ }\mu\text{m}$ [34].

Compared to the above MZ structure the femtosecond laser drilled microcavity MZIs exhibit a small size of the order of tens of microns and RI sensitivity comparable to that of nanocoated LPGs. The microcavity can be of different shapes as schematically shown in the Figure 5: conical [14,35,36], trench [37,38], circular [39–43].




Figure 5. Top view and cross-section of the three basic shapes of microcavities: (a) conical; (b) trench; (c) circular.

In all cases, the microcavity cuts partially through the core, so that part of the mode propagates in the core and another part in the cavity. The conical microcavity interferometer showed a RI sensitivity of about $9370 \text{ nm}/\text{RIU}$ for a diameter of $51 \text{ }\mu\text{m}$ [36], the trench cavity—about $10,800 \text{ nm}/\text{RIU}$ around 1.333 (water) and $-3243 \text{ nm}/\text{RIU}$ around 1 (air) for trench lengths from $50 \text{ }\mu\text{m}$ to $115 \text{ }\mu\text{m}$ [37,44]. The circular micro-cavity interferometer of diameter $60 \text{ }\mu\text{m}$ —about $12,000 \text{ nm}/\text{RIU}$ around 1.333 [38–40]. These and similar structures are referred as microcavity in-fiber MZI- μIMZI .

3.2. Structure of Micro-interferometers in Transmission and in Reflection

The μIMZI is a hole carved across the cladding of a single-mode fiber penetrating partially into the core of RI n_0 as shown in Figure 5a. The fundamental core mode propagating along the core has an approximately Gaussian distribution whose spread into the cladding is wavelength dependent, i.e., the higher the wavelength, the broader the distribution. When it reaches the cavity filled by a medium of lower RI (n), a part p ($-1 \leq p \leq 1$) of the mode electrical field is coupled into the cavity propagating at a different speed. For $p > 0$ the power is coupled from the core to the cavity while for $p < 0$ it is coupled back to the core. Since $n < n_0$ the beam is divergent, and losses occur [38]. At the other end of

the microcavity a relative power p couples back to the core after having accumulated a phase difference with respect to the part propagating in the core whose relative part of the electric field is q .

Power conservation demands that $p^2 + q^2 = 1$. In case of losses $p^2 + q^2 < 1$. The coupling process in case of transmission is schematically represented as shown in Figure 6a, while that in reflection is shown in Figure 6b. The coupling coefficient p increases with wavelength since mode field broadens. The structures in transmission are referred to as μ IMZI, while those in reflection as micro-cavity in-fiber reflective interferometer (μ IRI).




Figure 6. Microcavity interferometers: (a) core mode redistribution in a fiber with a microcavity; (b) mode coupling in a reflection scheme.

Responses and Sensitivities of Micro-Interferometers

Ignoring losses due to scattering and diffraction, after accumulating a phase delay the light is coupled back to the core. If the input intensity is I_0 , the intensity at the output of the μ IMZI in the case of no losses is defined as in Equation (3a) [36–40,44], while that of the μ IRI is as in Equation (3b), and the accumulated phase difference is as in Equation (3c).

$$I_T = I_1 + I_2 + 2\sqrt{I_1 I_2} \cos \Phi \quad I_1 = p^4 I_0, \quad I_2 = q^4 I_0 \quad (3a)$$

$$I_R = (I_1 + I_2 + 2\sqrt{I_1 I_2} \cos \Phi)^2 = I_T^2 \quad (3b)$$

$$\Phi = \frac{2\pi}{\lambda} \Delta n \cdot l + \phi \quad \Delta n = n_0 - n \quad (3c)$$

In Equation (3a,b) I_1 and I_2 are the intensities of the field in the core and the cavity respectively, $\Delta n = n_0 - n$ is the effective RI difference at wavelength λ between the core mode effective RI and that of the medium inside the cavity of effective length l_0 , ϕ being some constant phase shift accumulated because of microcavity shape imperfections.

A comparison between the responses of an interferometer in transmission and one in reflection shows that the losses in the reflective interferometer are twice higher in dB since from Equations (3a,b) $\log I_R = 2 \log I_T$. The advantage of a reflective microcavity interferometer is that it can be mounted, e.g., in a needle and be used as a tip sensor.

When a microcavity is drilled into the fiber, the exact positions of the minima cannot be fixed at will. However, for a practical device one would prefer to have the minimum at the right end of the spectral range to have sufficient spectral range for wavelength shifts caused by RI changes. To fine-tune the position of the minima extended microcavities can be used [38].

Both transmission and reflection structures show the same RI sensitivity [14,37] and free spectral range (FSR):

$$S_{\Delta n} = \frac{d\lambda}{d[\Delta n(\lambda)]} = \frac{d\lambda}{d(\Delta n)} = \frac{\lambda}{\Delta n(\lambda)} = -\frac{\lambda}{n_0(\lambda) - n(\lambda)} \quad (4a)$$

$$\Lambda = \frac{\bar{\lambda}^2}{\Delta n(\bar{\lambda}) l_0} = S_{\Delta n} \frac{\bar{\lambda}}{l_0}, \quad FSR = \frac{\bar{\lambda}^2}{\Delta n(\bar{\lambda}) l_0} = S_{\Delta n} \frac{\bar{\lambda}}{l_0} \quad \bar{\lambda} = \sqrt{\lambda_m \lambda_{m+1}} \quad (4b)$$

The sensitivity $S_{\Delta n}$ (nm/RIU) is thus proportional to the wavelength of the traced spectral minimum and inversely proportional to the effective modal RI difference while the length l_0 affects the free spectral range and not $S_{\Delta n}$. A μ IMZI and a μ IRI of same length l_0 will thus have the same sensitivity, but the minima of the latter will be twice deeper than those of the former.

3.3. Fabrication and Sensitivity Tuning of Micro in-Fiber Interferometers

3.3.1. Fabrication Technologies of μ IMZIs

Microcavity in-line Mach-Zehnder or reflective interferometers are usually fabricated using femtosecond lasers. The micromachining is similar to previously reported techniques [14,36–41,44] was performed using a Solstice Ti: Sapphire fs laser operating at $\lambda = 795$ nm as shown in Figure 7.




Figure 7. Schematic representation of the experimental set-up for the fabrication of a microcavity in a fiber using a femtosecond laser. A SEM view of the cross section of the microcavity.

The cavities are usually of size $d = 60$ μ m in diameter and $h_0 = 62$ μ m in depth fabricated in a SMF-28 fiber according to the procedure described in detail elsewhere [14,37–40]. In the experiments [39–43], the fiber was irradiated by 82 fs pulses at a repetition rate of 10 kHz. In order to make the microcavity, the laser beam is directed into a suitably designed micromachining setup based on the Newport μ Fab system. The system was equipped with a 20 \times lens, with NA = 0.30. Fiber transmission was monitored throughout the micromachining process with the NKT Photonics SuperK COMPACT supercontinuum white light source and Yokogawa AQ6370C optical spectrum analyzer in 1100–1700 nm spectral range. The fabrication process was controlled with an in-house developed software.

3.3.2. Fine-Tuning of μ IMZIs

Fine-tuning of the μ IMZI includes positioning the minimum to be tracked at a desirable wavelength and increasing the sensitivity. The first approach is reported in [38] where it is proposed to machine an extension of the microcavity and by changing its width and depth a minimum can be tuned by 200 nm and the sensitivity increased by 32% reaching 26,880 nm/RIU in the 1.3333–1.3412 RIU. The second approach is to apply reactive ion etching in oxygen plasma [39]. Etching improves wettability and thus improves sensitivity from 11,377 nm/RIU to 11,800 nm/RIU around 1.333.

Thin Al_2O_3 film deposition gives the possibility [41] for a transition from volume to surface sensitivities, which gives the possibility for responses when surface RI changes due to the presence of bacteria, for example. The above-cited values for the sensitivities are measured in the 1400 nm to 1700 nm range and the insertion losses are around 10 dB. According to Equation (4a) sensitivity $S_{\Delta n}$ around 800 nm will be twice lower when compared to that at 1600 nm, which means that it would be of the order of 6000 nm/RIU. Since the fibers to be used must be single mode at 800 nm, the insertion losses would increase because of the smaller radius. However, tapering the fiber and drilling the microcavity in the tapered part decreases losses down to 5 dB and increases sensitivity up to 14,000 nm/RIU for SM800 fiber and up to 11,500 nm/RIU in SM600 nm [42].

One general advantage of microcavity interferometers is the extremely small amount of sensing volumes. For cavity diameters from $d = 50 \mu\text{m}$ to $60 \mu\text{m}$ and a depth of $h = 60 \mu\text{m}$ the volume is from $V = 117 \text{ pL}$ to $V = 170 \text{ pL}$. Using nanocoating on the bottom for a surface sensitivity means that the outer cladding of the fiber can be reduced in a way that the microcavity depth reduces to about $10 \mu\text{m}$ which would reduce the sensing volume to less than 50 pL .

Table 1 summarizes the RI sensitivities of various optical fiber sensing platforms considered above. As seen, standard LPGs and interferometric structures cannot reach sensitivity above 2000 nm/RIU except for a sensor using evanescent field in an asymmetrical taper. LPGs around DTP can reach thousands of nm/RIU and additional etching and nano-coating can increase it to $20,000 \text{ nm/RIU}$. The μIMZIs on the other hand, usually exhibit sensitivity from 9000 to $11,000 \text{ nm/RIU}$ without additional processing or coating. Chemical etching and cavity extension can increase sensitivities to record values of $17,200$ and $27,000 \text{ nm/RIU}$.

Table 1. Comparison of Mach-Zehnder interferometers, LPGs, and in-line Mach-Zehnder interferometers (μIMZI), and their sensitivities around water RI of 1.333.

Structure	Specific Features	$S_{\Delta n}$ [nm/RIU]	Reference
LPG	Double resonance	750	[10]
LPG	SiO_2 nanoparticle mesoporous thin films	1927	[9]
Twin-core	SMF-Twin core-SMF	1973	[17]
D-shaped	Nanocoated D-shaped fiber	1566	[19]
SMS	fs induced spiral MSF	2144	[21]
MZI	PCF	100–150	[16]
MZI	Etched PCF	360	[34]
Tapered	SMF	415	[31]
Tapered	Multicore fiber	703	[33]
Taped fiber	Asymmetrical taper evanescent field	3390–3914	[18]
DTP-LPG	Cladding etching and nano-coating	20,000	[26]
μIMZI	Conical	9370	[35]
μIMZI	Conical	10,537	[14]
μIMZI	Conical and chemical etching	17,197	[36]
μIMZI	Plasma-processed	11,800	[39]
μIMZI	Trench	10,000	[44]
μIMZI	Trench	10,780	[37]
μIMZI	Circular, thin film deposition	12,390	[41]
μIMZI	Circular, extended	26,882	[38]
μIRI	In reflection	13,865	[40]
μIMZI	Vis/NIR tapered	12,000	[42]

4. Bacteria Detection Using LPGs

In literature, throughout the years we can find several LPG-based bacteria detection systems. They differ in the design of the sensing part, the choice of the bioreceptor, and surface functionalization. However, each fiber optic biosensor essentially comprises a biorecognition element—a bioreceptor which drives the selectivity and specificity of the sensor, chemical functionalization to immobilize the bioreceptor, and properly orient it on the sensor's surface, and a transducer able to convert the target-receptor binding into a measurable and detectable signal (Figure 8).

In our work, we will divide the sensors based on the target—whether the described method aim detection of whole-cell (direct) or just a part of bacteria (indirect)—and in terms of the biorecognition element.




Figure 8. Schematic representation of the fiber optic biosensor.

4.1. Direct Methods

Direct methods refer to the detection of whole bacteria cells. When early and fast diagnosis is required, this system is one of the best ways to identify the pathogenic strain of bacteria. The main advantages of this approach are toleration of the whole-cells to the environmental changes such as temperature, pH, etc. in contrast to isolated structures; the ease of the isolation of the bacteria cells from natural sources; and no need for extensive preparation of the samples before the measurement process. Nonetheless, the development of the sensing system for the whole bacteria cell is challenging. We need to take into consideration the size of the bacteria, which is much larger than other analytes, such as proteins, DNA aptamers, and its surface structures, e.g., pili and flagella, which can disturb measurements. Moreover, a number of the outer structures on the bacteria surface can serve as biorecognition/binding places and may lead to nonspecific interactions with the sensor's surface. Moreover, depending on the form, which we target to detect, whether it will be live bacteria, dry mass, or heat-killed organism, they will act differently during the assay. Thus, to avoid the false-positive interactions the biorecognition element on the sensor's surface supposed to be well chosen, as it will be responsible for specificity and partially for the efficiency of the whole experiment.

4.1.1. Bacteriophages

The first published attempt to bacteria detection using LPG was done utilizing the T4 bacteriophages as a biorecognition element. As it is a natural and very specific enemy of the *E. coli* bacteria, it could be successfully introduced to the whole-cell bacteria detection. In [43], Smietana et al. reported the LPG with the sensitivity ~ 600 nm/RIU and physically adsorbed bacteriophages on the LPG surface. Due to the relatively low sensitivity of the platform, proof-of-concept character of the studies, and the non-covalent link between the surface and the bioreceptor, only high concentrations of *E. coli* were detected, with no connection between the targeted bacteria concentration and obtained spectral shift.

One year later, Tripathi et al. [45] reported the same approach, however, to assure the covalent binding of the T4 phages the chemical functionalization of the surface was introduced. The silanization with APTES and activation of the obtained $-\text{NH}_2$ groups using glutaraldehyde provided $-\text{COOH}$ groups, which could be easily bonded to the $-\text{NH}_2$ groups present on the T4. As the sensor had higher sensitivity reaching ~ 2000 nm/RIU, and the bioreceptor was covalently bound to the sensor's surface, the sensing schema was tested against different concentrations of *E. coli* bacteria in colony-forming units (CFU) and resulted in 10^3 CFU/mL limit of the detection.

Although the T4 bacteriophages are very specific, they also have limitations. Considering highly complex structure, they are prone to mechanical damages and, what is more, even after chemical functionalization of the LPG's surface we cannot be sure about their orientation on the sensor's surface. In addition, the T4 is the lytic phage, which means that after binding the bacteria cell and infection, the *E. coli* cells are destroyed releasing newly replicated phages, which, in turn, can lead to false-negative/positive signals.

In response to that, Chiniforooshan et al. [46] reported a comparison of the *E. coli* detection using T4 and MS2 phages. The difference in shape, robustness, place of the attachment, the infection cycle

of the MS2 and surface functionalization performed utilizing gas instead of liquid phase resulted in lowering the limit of the detection from 10^3 to 100 CFU and higher stability of the measurements. The sensitivity of the LPG utilized during this experiment reached ~ 7000 nm/RIU in the RI range from 1.33–1.34 RIU.

4.1.2. Adhesins—Receptor Binding Proteins

The relatively large size of bacteriophages and drawbacks associated with them can be a problem for some biosensing applications. Therefore, in some cases, just receptor binding proteins located on the phages' fiber tails can be applied. These proteins are responsible for specific recognition of host-bacteria and one of their biggest advantages is the possibility of introducing changes to their sequence/structure [47]. Without disturbing and altering their binding affinity, we can improve protein's orientation on the sensor's surface [48]. The adhesin utilized on LPG for label-free *E. coli* bacteria detection is protein gp37 obtained from phage T4.

Brzozowska et al. [49] and Koba et al. [50] reported adhesin's gp37 ability to detect whole bacteria cells and presented its specificity. In both cases, for the experiment LPGs with RI sensitivity reaching ~ 3000 nm/RIU have been utilized. The functionalization process applied in these experiments was intended for proteins containing histidine (His) tag. The chemical functionalization of LPG included silanization of the surface in GPTMS solution. Next, the LPG was immersed in 20 mM (millimolar) iminodiacetic acid in NaHCO_3 buffer. The last step included the immobilization of the nickel ions on the sensor's surface. Such prepared surface was ready to attach the bioreceptor—adhesin gp37. The goal of these studies was to determine the sensor's specificity; thus, it was verified with many bacterial strains such as *S. enterica*, *K. pneumoniae*, *E. coli* K12, or *Lactobacillus* 919. All the strains were detected in one concentration—namely 100 $\mu\text{g}/\text{mL}$ as the lowest limit of detection (LOD) was not the priority. The research has clearly proven the possibility of using adhesins in place of bacteriophages for specific bacteria strain recognition.

Next work utilizing the adhesin gp37 as a bioreceptor was reported by Piestrzynska et al. [51]. Here, a label-free biosensing application of LPGs optimized in RI sensitivity by deposition of thin tantalum oxide (TaO_x) overlays was discussed. The TaO_x nano-coated LPGs utilized during the experiment showed extremely high RI sensitivity $\sim 11,500$ nm/RIU in RI range 1.335–1.345 RIU. Besides enhancing the RI sensitivity, TaO_x was also expected to provide the chemical and mechanical stability during the surface functionalization or biosensing procedure. For the *E. coli* bacteria sensing experiment, the TaO_x nano-coated LPG was first subjected to surface silanization with triethoxysilylpropyl succinic anhydride (TESPSA). Next, the formation of an amide bond between amine groups in the Ni-NTA complex and groups of succinic anhydrides on the surface of the fiber enabled attachment of the adhesin containing His-tag at the N-terminus. Despite highly increased initial sensitivity of the nanocoated LPGs, the LOD reached just 10^3 CFU/mL. The experiment proved that during the work with such sensitive sensors, the functionalization process, along with the size of the bioreceptor and the biological target is crucial. Thus, every modification of such LPG needs to be taken into consideration because it may significantly reduce initially optimized RI sensitivity.

4.1.3. Antibodies

The next group of the LPG-based sensors includes devices utilizing antibodies (Abs) as bioreceptors. Abs are also known as immunoglobulins (Ig), are large, “Y-shape” proteins with dimensions approximately $14 \times 10 \times 4$ nm [52]. Each Ab molecule has its unique structure that enables it to bind specifically to its corresponding target. They have several advantages such as high specificity, versatility and thanks to the ease of integration into different systems, Abs were combined with numerous sensor formats. Once they are selected and secreted, the immobilization process becomes crucial for their efficiency. The control over the position and orientation of Abs during the covalent binding is challenging. In an ideal scenario, antibodies should be immobilized in their native form and oriented to maximize complementary binding, without the need for introducing functional

groups or special tags/labeling. To date, no method genuinely provides the ideal scenario [53]. Moreover, depending on the target, Abs can provide different immobilization efficiency of the detected pathogen [54] and they are prone to physical, chemical, and enzymatic damages. However, despite mentioned drawbacks, so-called immunosensors have been regarded as a gold standard in diagnostics and environmental monitoring.

The use of Abs was selected for the detection of bacteria like *E. coli* and *S aureus*. In the detection of *S. aureus*, Bandara et al. [55] functionalized the LPG surface using layer-by-layer (LbL) deposition of self-assembled polyelectrolyte multilayers. Immersion of the fiber in PCBS resulted in terminal carboxyl groups, which were then covalently conjugated to the Abs. Chosen Abs were specific to penicillin-binding-protein 2a of methicillin-resistant (MR) staphylococci. The binding of the target bacteria to the sensor was detected as a change in the transmitted power through the fiber at the wavelength of peak attenuation of the turning around point-LPG (1550 nm). In a buffer solution, a LOD of 100 CFU was reached, but they also showed the selectivity of the sensor testing 36 different strains of MR staphylococci (all at 10^4 CFU concentration).

The same approach in terms of LPG's functionalization was also implemented by Fang et al. [56] in the microfluidic-LPG schema. However, in this work polyelectrolyte, functional coatings PAH/PAA were additionally modified to selectively facilitate the bacterial adhesion on the surfaces. Moreover, the functional coatings not only increase the initial rate of bacterial adhesion onto the surfaces (as the experiment was conducted in the constant flow), but also improve the sensitivity of LPG and thus the detection limit for *S. aureus* (calculated 224 CFU/mL, experimental range 10^4 – 10^8 CFU).

Another configuration of LPG used in immunosensing presented by Kaushik et al. [57] is a combination of two identical chirped LPGs for *E. coli* detection. The described sensor with an inter-grating space works like a Mach-Zehnder interferometer. To immobilize monoclonal Abs the surface was silanized with APTES to generate the $-NH_2$ groups and were activated with EDC/NHS. This results in the activation of carboxyl groups of anti-*E. coli* antibodies to form covalent bonding with the amine groups on the sensing probe via carbodiimide cross linker chemistry. Such prepared immunosensor was then exposed to different concentrations of *E. coli* (10 CFU/MI–60 CFU/mL) and the obtained spectral shift for the lowest concentration reached 0.2 nm for 10 CFU/mL (calculated LOD 7 CFU).

4.1.4. Carbohydrates

Another and a relatively new approach to bacteria detection in a view of bioreceptor is the use of carbohydrates. This system is based on the natural affinity of bacteria to specific carbohydrates, a moiety occurring in almost every living organism, which is a basis of the bacteria-host cell recognition. Therefore, they are able to naturally attract bacteria, facilitate biofilm formation, and promote bacteria adhesion [24]. In this sensing schema, the detection of pathogenic bacteria is carried out via protein-carbohydrate interactions, which are very similar to the antibody-antigen, or enzyme-substrate reactions.

Celebanska et al. [58] reported the LPFG-based system for *E. coli* detection. During the experiment, highly sensitive ~ 7000 nm/RIU LPG was used. The fiber optic platform was functionalized with APTES by a gas phase method. Next, the amine groups were activated by a homobifunctional cross linker PDITC to covalently bind the chosen carbohydrate— α -mannose (APMAN)—on the sensor's surface. The sensor was examined with specific *E. coli* bacteria within the range 10^3 to 10^8 CFU/mL. Despite the very high sensitivity of the LPFG platform, the experimental LOD reached just 10^3 CFU. Moreover, the much broader specificity in comparison to the bacteriophages or antibodies disqualifies the carbohydrates from one strain-targeted tests; however, such combination could be used as an entry test for potentially contaminated samples.

4.2. Indirect Methods

Indirect methods refer to the detection of chosen parts of bacteria. Depending on the type of bacteria, we can target different surface structures such as lipopolysaccharides (LPSs), outer membrane proteins (OMPs), porins, or glycoproteins. Even though they need isolation and preparation before the use, they are a very good starting point for initial tests, which are in most cases the base for further phases targeting the detection of the whole microorganisms. What is more, it is a very good solution for laboratories without special facilities needed to handle the work with whole bacteria (especially pathogenic strains). To date, three groups reported the indirect detection of *E. coli* using an LPG-based system.

4.2.1. Adhesins—Receptor Binding Proteins

The first work describes the use of the phage T4 adhesin gp37 mentioned in Section 4.1.2. Brzowska et al. [58] showed the detection of *E. coli* LPS using a highly sensitive platform (3000 nm/RIU). The adhesin was again covalently bound to the surface by histidine-tag using nickel ions and was tested in LPS concentration equal to 0.2 mg/mL. Koba et al. [59] utilized the same approach to show the stability and reusability of the presented sensing system. Both presented works were just an introduction to use the adhesin as a bioreceptor for whole-cell bacteria detection, which was described in later articles.

4.2.2. DNA Aptamers

Indirect detection was also described in the configuration where DNA aptamers were employed as a bioreceptor. An aptamer is a short (25 to 90 bases [60]), single-stranded nucleic acid sequence DNA or RNA, which can attach to a variety of targets like bacteria, viruses, proteins, drugs, or even heavy metal ions. Thanks to their three-dimensional structure, the aptamers can bind the target with high efficacy and specificity—comparable to, or even higher than Abs, what may be caused by aptamer's capability of folding upon binding target [61]. They provide exceptional flexibility and convenience regarding the engineering of their structure what have led to the development of DNA aptamers with enhanced affinity, specificity, and stability [62]. In comparison to Abs, aptamers are structurally stable across a wide range of temperatures and storage conditions. Due to their outstanding binding characteristics, stability and possibility for easy modifications, aptamers have all the necessary qualities to be used as the basis of detection elements. Surprisingly, despite the availability of aptamers recognizing a wide range of pathogens, a relatively low number of them have yet been used in biosensing.

Queirós et al. [63] presented an LPG-based aptasensor for selective *E. coli* outer membrane protein detection. To covalently bind the DNA, the LPGs surface was functionalized using two different processes: an electrostatic method, using PLL as cationic polymer and a covalent method, using APTES. The LPG with the sensitivity of ~150 nm/RIU was tested with known protein concentration equal to 0.1 to 10 nm. In the end, both functionalization methods gave an identical performance in terms of detection; however, polycationic functionalization presented better results in terms of reproducibility and uniformity of functionalization.

5. Bacteria Detection Using μ IMZI

LPGs were tested throughout the years also for the detection of many different bio-applications. Although they appeared effective and reached very low limits of detection, many challenges remain valid and urgent for the creation of successful biosensing systems. Common limiting factors of the LPGs' are their capability for multiplexing or reproducibility, especially more complex platforms. Moreover, this type of sensor having 4–5 cm length of the fiber usually requires at least 500 μ L of the examined sample, which can be problematic when only small volumes of analyte are available. Most of the LPG sensors are also sensitive towards temperature, which is an essential disadvantage since the temperature is one of the most disrupting factors in RI measurements. Despite generally good

performance, there are still drawbacks and open questions about the most precise and reliable detection. Identification of small molecules in *pl* volumes—possibly up to single-molecule—undisturbed by external factors, along with miniaturization of the device is still a challenge. Therefore, constant attention is paid to developing new, cutting-edge devices that will be capable of solving these challenges.

One of the relatively new classes of sensing devices are structures based on in-fiber microcavities [15]. The presence of tens of micrometers diameter cavity micromachined in the fiber is the defining feature of these platforms. Although the microcavity-based sensors are known for almost a decade and have been reported by few groups in different shapes and for different applications, the subject of using it for specific biosensing has been reported just once so far. The only work using this sensing scheme for specific bacteria detection up to date was published by Janik et al. [64]. Microcavity in-line Mach-Zehnder interferometer with 60 μm diameter and sensitivity reaching 15,000 nm/RIU was functionalized with APTES from a gas phase assuring uniform chemical layer on the sensor's surface. Activated by GLU $-\text{NH}_2$ groups were able to covalently attach chosen bioreceptor—MS2 bacteriophages. As described in Section 3, due to their properties, they evenly cover the surface of the microcavity. Such a prepared sensor was tested with different *E. coli* bacteria concentrations ranging from 100 to 10^8 CFU and was able to detect 100 CFU of live bacteria in liquid volume as low as picoliters. Considering novel and miniature design, superior sensitivity, unique feature to measure volumes as low as 117–170 picoliters as shown above, temperature insensitivity, an in-fiber microcavity structure, and a fiber tip design would be a good alternative with a very high potential in biosensing.

Table 2 below summarizes the reported *S. aureus* and *E. coli* bacteria sensing results with LPGs and μIMZI .

Table 2. Fiber optic biosensors for bacteria detection.

Sensor Type	Sensitivity of the Sensor [nm/RIU]	Type of Detected Bacteria	Range of the Detection [CFU/mL]	Surface Functionalization	Receptor	Ref
DIRECT METHODS						
Chirped-LPG	-	<i>E. coli</i>	10–60 CFU/mL	APTES+EDC	Ab	[65]
LPG	7200	<i>E. coli</i>	10^3 CFU/mL	APTES+PDITC	APMAN	[58]
2 LPG	1929	<i>E. coli</i>	10^2 – 10^5 CFU/mL	APTES + GLU	T4 phages	[66]
LPG	570	<i>E. coli</i>	10^8 CFU/mL	-	T4 phages	[45]
LPG	-	<i>E. coli</i>	100 CFU/mL	APTES+GLU	MS2 phages	[47]
LPG	-	<i>S. aureus</i>	10^4 – 10^8 CFU/mL	ISAM ([PAH/PAA] ₁₀ PAH)	Ab	[56]
LPG	-	<i>S. aureus</i>	10^2 – 10^6 CFU/mL	ISAM (PCBS)	Ab	[55]
LPG	2600	<i>E. coli</i>	100 $\mu\text{g/mL}$	GPTMS+Ni + His-Tag	adhesin gp37	[50]
LPG	3000	<i>E. coli</i>	100 $\mu\text{g/mL}$	GPTMS +Ni + His-Tag	adhesin gp37	[51]
LPG	2321	<i>E. coli</i>	10^3 – 10^9 CFU/mL	APTES+GLU	T4 phages	[46]
LPG	11,000	<i>E. coli</i>	10^3 – 10^9 CFU/mL	TESPSA+Ni + His-Tag	gp37	[52]
μIMZI	15,000	<i>E. coli</i>	10^2 – 10^8 CFU/mL	APTES+GLU	MS2 phages	[63]
INDIRECT METHODS						
LPG	6947.4	LPS <i>E. coli</i> B	0.2 mg/mL	GPTMS+Ni + His-Tag	adhesin gp37	[59]
LPG	2600	LPS <i>E. coli</i> B	0.2 mg/mL	GPTMS+Ni + His-Tag	adhesin gp37	[67]
LPG	150	OMPs <i>E. coli</i>	0.1–10 nM	ISAM (PLL) APTES	DNA aptamer	[63]

6. Conclusions

In this paper, we presented the principle of operation of two highly RI sensitive platforms, i.e., long-period gratings operating around dispersion turning point (DTP-LPGs) and microcavity in-line Mach-Zehnder interferometer (μ IMZI). Different types of Mach-Zehnder interferometers and of LPGs have been reviewed and their sensitivities have been compared. Comparison shows that in terms of sensitivity to RI the most promising structures are DTP-LPGs and μ IMZIs with typical sensitivities around water RI from 2600 nm/RIU to 11,000 nm/RIU and around 10,000 nm/RIU for μ IMZI. With additional fine-tuning procedures, these sensitivities can be increased to correspondingly 20,000 nm/RIU and 27,000 nm/RIU. However, while sensitivity of DTP-LPGs reduces thereafter, that of μ IMZI increases. Moreover, so far, DTP-LPGs have been demonstrated in a transmission mode, while microcavity MZIs have been demonstrated in reflection mode with a sensitivity of 12,000 nm/RIU.

Up to now, the DTP-LPGs have been more intensively explored as fiber-based sensing platforms for direct and indirect *E. coli* and *S. aureus* bacteria detection. A variety of surface functionalizations with different receptors as bacteriophages, adhesins, antibodies, and carbohydrates, and DNA aptamers, have been tested. Ranges over four orders of magnitude (10^2 – 10^6 CFU/mL and 10^4 – 10^8 CFU/mL) for *S. aureus* detection and over six orders of magnitude (10^3 – 10^9 CFU/mL) for *E. coli* have been demonstrated. Only one report so far has demonstrated *E. coli* detection of six orders of magnitudes (10^2 – 10^6 CFU/mL) using μ IMZI. Taking into account the higher sensitivity of μ MZIs and the possibility to fabricate them in reflection mode at the tip of a fiber, makes them promising for highly sensitive sub-nano- and picoliter dip biosensors mounted in medical needles.

Author Contributions: Conceptualization, T.E. and M.J.; writing—original draft preparation, T.E. and M.J.; writing—review and editing, T.E., M.J., M.K. and M.Ś.; supervision, W.B., M.K. and M.Ś.; project administration M.Ś.; design and implementation of the fs laser microfabrication schemes, M.J. and M.K.; software: data acquisition and measurement algorithms, M.K.; P.M. fabricated part of the samples (LPGs and micro IMZIs) and maintains the femtosecond laser; funding acquisition, W.B. and M.Ś. All authors have read and agreed to the published version of the manuscript.

Funding: This work was supported by the National Science Center, Poland (NCN) (2018/29/B/ST7/02552).

Conflicts of Interest: The authors declare no conflict of interest.

References

1. CDC. Estimates of Foodborne Illness in the United States Centers for Disease Control and Prevention. 2011. Available online: http://www.cdc.gov/foodborneburden/PDFs/FACTSHEET_A_FINDINGS_updated4-13.pdf (accessed on 15 June 2020).
2. Delbeke, S.; Ceuppens, S.; Holvoet, K.; Samuels, E.; Sampers, I.; Uyttendaele, M. Multiplex real-time PCR and culture methods for detection of Shiga toxin-producing *Escherichia coli* and *Salmonella* Thompson in strawberries, a lettuce mix and basil. *Int. J. Food Microbiol.* **2015**, *193*, 1–7. [[CrossRef](#)] [[PubMed](#)]
3. Law, J.W.-F.; Mutalib, N.-S.A.; Chan, K.-G.; Lee, L.-H. Rapid methods for the detection of foodborne bacterial pathogens: Principles, applications, advantages and limitations. *Front. Microbiol.* **2015**, *5*, 770. [[CrossRef](#)] [[PubMed](#)]
4. Silva, D.M.; Domingues, L. On the track for an efficient detection of *Escherichia coli* in water: A review on PCR-based methods. *Ecotoxicol. Environ. Saf.* **2015**, *113*, 400–411. [[CrossRef](#)] [[PubMed](#)]
5. James, S.W.; Tatam, R.P. Optical fibre long-period grating sensors: Characteristics and application. *Meas. Sci. Technol.* **2003**, *14*, R49–R61. [[CrossRef](#)]
6. Rego, G. A Review of Refractometric Sensors Based on Long Period Fibre Gratings. *Sci. World J.* **2013**, *2013*, 1–14. [[CrossRef](#)]
7. Petrovic, J.S.; Dobb, H.; Mezentsev, V.K.; Kalli, K.; Webb, D.J.; Bennion, I. Sensitivity of LPGs in PCFs Fabricated by an Electric Arc to Temperature, Strain, and External Refractive Index. *J. Lightwave Technol.* **2007**, *25*, 1306–1312. [[CrossRef](#)]
8. Li, L.; Xia, L.; Xie, Z.; Liu, D. All-fiber Mach-Zehnder interferometers for sensing applications. *Opt. Express* **2012**, *20*, 11109–11120. [[CrossRef](#)]

9. Korposh, S.; Lee, S.-W.; James, S.W.; Tatam, P. Refractive index sensitivity of fibre-optic long period gratings coated with SiO₂nanoparticle mesoporous thin films. *Meas. Sci. Technol.* **2011**, *22*, 075208. [[CrossRef](#)]
10. Li, Q.-S.; Zhang, X.-L.; Shi, J.-G.; Xiang, D.; Zheng, L.; Yang, Y.; Yang, J.-H.; Feng, D.; Dong, W.-F. An Ultrasensitive Long-Period Fiber Grating-Based Refractive Index Sensor with Long Wavelengths. *Sensors* **2016**, *16*, 2205. [[CrossRef](#)]
11. Gambhir, M.; Gupta, S. Review of Turn around Point Long Period Fiber Gratings. *J. Sens. Technol.* **2015**, *5*, 81–89. [[CrossRef](#)]
12. Chiavaioli, F.; Baldini, F.; Tombelli, S.; Trono, C.; Giannetti, A. Biosensing with optical fiber gratings. *Nanophotonics* **2017**, *6*, 663–679. [[CrossRef](#)]
13. Dyankov, G.; Eftimov, T.; Malinowski, N.; Belina, E.; Kisov, H.; Mikulic, P.; Bock, W. A highly efficient biosensor based on MAPLE deposited hemoglobin on LPGs around phase matching turning point. *Opt. Laser Technol.* **2020**, *123*, 105907. [[CrossRef](#)]
14. Zhao, Y.; Zhao, H.; Lv, R.-Q.; Zhao, J. Review of optical fiber Mach–Zehnder interferometers with micro-cavity fabricated by femtosecond laser and sensing applications. *Opt. Lasers Eng.* **2019**, *117*, 7–20. [[CrossRef](#)]
15. Park, M.; Lee, S.; Ha, W.; Kim, D.-K.; Shin, W.; Sohn, I.-B.; Oh, K. Ultracompact Intrinsic Micro Air-Cavity Fiber Mach–Zehnder Interferometer. *IEEE Photonics Technol. Lett.* **2009**, *21*, 1027–1029. [[CrossRef](#)]
16. Wang, J.-N.; Tang, J.-L. Photonic Crystal Fiber Mach-Zehnder Interferometer for Refractive Index Sensing. *Sensors* **2012**, *12*, 2983–2995. [[CrossRef](#)]
17. Guzman-Sepulveda, J.R.; Lopez-Cortes, D.; Romano, I.H.; Margulis, W.; May-Arrijoja, D.A. Refractive Index Sensor using a Two-Core Optical Fiber. In Proceedings of the Conference on Lasers and Electro-Optics, San Jose, CA, USA, 6–11 May 2012.
18. Kamil, Y.M.; Mustapa, M.A.; Abu Bakar, M.; Musa, B.; Syahir, A.; Yaacob, M.H.; Mahdi, M.; Mustapha, Y.K. Refractive index sensor with asymmetrical tapered fiber based on evanescent field sensing. In Proceedings of the 2015 IEEE International Broadband and Photonics Conference (IBP), Bali, Indonesia, 23–25 April 2015; pp. 35–39.
19. Del Villar, I.; Zubiarte, P.; Zamarreño, C.R.; Arregui, F.J.; Matias, I.R. Optimization in nanocoated D-shaped optical fiber sensors. *Opt. Express* **2017**, *25*, 10743. [[CrossRef](#)]
20. Chen, C.-H.; Tsao, T.-C.; Tang, J.-L.; Wu, W.-T. A Multi-D-Shaped Optical Fiber for Refractive Index Sensing. *Sensors* **2010**, *10*, 4794–4804. [[CrossRef](#)]
21. Liu, F.; Lin, H.F.; Liu, Y.; Zhou, A.; Dai, Y.T. Femtosecond-induced spiral micro-structured SMS fiber structure for refractive index measurement. *Opt. Express* **2018**, *26*, 17388. [[CrossRef](#)]
22. Erdogan, T. Fiber grating spectra. *J. Lightwave Technol.* **1997**, *15*, 1277–1294. [[CrossRef](#)]
23. Rego, G. Arc-Induced Long Period Fiber Gratings. *J. Sens.* **2016**, *2016*, 1–14. [[CrossRef](#)]
24. Schlangen, S.; Bremer, K.; Zheng, Y.; Böhm, S.; Steinke, M.; Wellmann, F.; Neumann, J.; Roth, B.W.; Overmeyer, L. Long-Period Gratings in Highly Germanium-Doped, Single-Mode Optical Fibers for Sensing Applications. *Sensors* **2018**, *18*, 1363. [[CrossRef](#)] [[PubMed](#)]
25. Smietana, M.; Bock, W.J.; Mikulic, P. Comparative study of long-period gratings written in a boron co-doped fiber by an electric arc and UV irradiation. *Meas. Sci. Technol.* **2010**, *21*, 25309. [[CrossRef](#)]
26. Smietana, M.; Koba, M.; Mikulic, P.; Bock, W.J. Towards refractive index sensitivity of long-period gratings at level of tens of μm per refractive index unit: Fiber cladding etching and nano-coating deposition. *Opt. Express* **2016**, *24*, 11897. [[CrossRef](#)] [[PubMed](#)]
27. Smietana, M.; Myśliwiec, M.; Mikulic, P.; Witkowski, B.S.; Bock, W.J. Capability for Fine Tuning of the Refractive Index Sensing Properties of Long-Period Gratings by Atomic Layer Deposited Al₂O₃ Overlays. *Sensors* **2013**, *13*, 16372–16383. [[CrossRef](#)]
28. Carter, R.M.; Maier, R.R.J.; Biswas, P.; Basumallick, N.; Bandyopadhyay, S.; Jones, B.J.S.; McCulloch, S.; Barton, J.S.; Bandyopadhyay, S. Experimental Difficulties with LPG Sensors Operating close to the Phase Turning Points. *J. Lightwave Technol.* **2016**, *34*, 3999–4004. [[CrossRef](#)]
29. Tripathi, S.M.; Bock, W.J.; Kumar, A.; Mikulic, P. Temperature insensitive high-precision refractive-index sensor using two concatenated dual-resonance long-period gratings. *Opt. Lett.* **2013**, *38*, 1666–1668. [[CrossRef](#)] [[PubMed](#)]
30. Park, M.; Lee, S.; Ha, W.; Kim, D.-K.; Shin, W.; Sohn, I.-B.; Oh, K. Fabrication of Intrinsic Fiber Mach-Zehnder Interferometer by Imbedding Micro Air-cavity. In Proceedings of the Conference on Lasers and Electro-Optics, Baltimore, MD, USA, 31 May–5 June 2009.

31. Ahsani, V.; Ahmed, F.; Jun, M.B.; Bradley, C. Tapered Fiber-Optic Mach-Zehnder Interferometer for Ultra-High Sensitivity Measurement of Refractive Index. *Sensors* **2019**, *19*, 1652. [[CrossRef](#)]
32. Han, X.; Liu, C.; Jiang, S.; Leng, S.; Yang, J. Up-down Taper Based In-Fiber Mach-Zehnder Interferometer for Liquid Refractive Index Sensing. *Sensors* **2019**, *19*, 5440. [[CrossRef](#)]
33. Abbas, Z.; Mashhadani, A.; Navruz, I. Ultrahigh sensitive refractive index sensor based on tapered multicore optical fiber. *Commun. Fac. Sci. Univ. Ankara* **2018**, *60*, 1–14.
34. Du, H.; Sun, X.; Hu, Y.; Dong, X.; Zhou, J. High Sensitive Refractive Index Sensor Based on Cladding Etched Photonic Crystal Fiber Mach-Zehnder Interferometer. *Photonic Sens.* **2019**, *9*, 126–134. [[CrossRef](#)]
35. Wang, Y.; Yang, M.; Wang, D.N.; Liu, S.; Lu, P. Fiber in-line Mach-Zehnder interferometer fabricated by femtosecond laser micromachining for refractive index measurement with high sensitivity. *J. Opt. Soc. Am. B* **2010**, *27*, 370. [[CrossRef](#)]
36. Sun, X.-Y.; Chu, D.-K.; Dong, X.-R.; Zhou, C.; Li, H.-T.; Zhi, L.; Hu, Y.-W.; Zhou, J.-Y.; Wang, C.; Duan, J.-A. Highly sensitive refractive index fiber inline Mach-Zehnder interferometer fabricated by femtosecond laser micromachining and chemical etching. *Opt. Laser Technol.* **2016**, *77*, 11–15. [[CrossRef](#)]
37. Jiang, L.; Zhao, L.; Wang, S.; Yang, J.; Xiao, H. Femtosecond laser fabricated all-optical fiber sensors with ultrahigh refractive index sensitivity: Modeling and experiment. *Opt. Express* **2011**, *19*, 17591–17598. [[CrossRef](#)] [[PubMed](#)]
38. Janik, M.; Eftimov, T.; Koba, M.; Smietana, M.; Bock, W.J. Tailoring Properties of Microcavity In-Line Mach-Zehnder Interferometer by the Microcavity Enlargement Using Femtosecond Laser. *J. Lightwave Technol.* **2019**, *37*, 4501–4506. [[CrossRef](#)]
39. Janik, M.; Mysliwiec, A.K.; Koba, M.; Celebanska, A.; Bock, W.J.; Smietana, M. Sensitivity Pattern of Femtosecond Laser Micromachined and Plasma-Processed In-Fiber Mach-Zehnder Interferometers, as Applied to Small-Scale Refractive Index Sensing. *IEEE Sens. J.* **2017**, *17*, 3316–3322. [[CrossRef](#)]
40. Eftimov, T.; Mikulic, P.; Arapova, A.; Bock, W.J. Ultrahigh refractive index sensitivity with a reflective fibre optic microcavity in-line interferometer. In Proceedings of the 27th International Conference on Optical Fiber Sensors, Alexandria, VA, USA, 8–12 June 2020.
41. Bock, W.J. Transition between bulk and surface refractive index sensitivity of micro-cavity in-line Mach-Zehnder interferometer induced by thin film deposition. *Opt. Express* **2017**, *25*, 26118. [[CrossRef](#)]
42. Eftimov, T.; Janik, M.; Bock, W.J. Microcavity In-Line Mach-Zehnder Interferometers Fabricated in Single-Mode Fibers and Fiber Tapers for Visible (VIS) and Near-Infrared (NIR) Operation. *J. Lightwave Technol.* **2019**, *37*, 3351–3356. [[CrossRef](#)]
43. Smietana, M.; Bock, W.J.; Mikulic, P.; Ng, A.; Chinnappan, R.; Zourob, M. Detection of bacteria using bacteriophages as recognition elements immobilized on long-period fiber gratings. *Opt. Express* **2011**, *19*, 7971–7978. [[CrossRef](#)]
44. Zhao, L.; Jiang, L.; Wang, S.; Xiao, H.; Guo, L.B.; Tsai, H.-L. A High-Quality Mach-Zehnder Interferometer Fiber Sensor by Femtosecond Laser One-Step Processing. *Sensors* **2010**, *11*, 54–61. [[CrossRef](#)]
45. Tripathi, S.M.; Bock, W.J.; Mikulic, P.; Chinnappan, R.; Ng, A.; Tolba, M.; Zourob, M. Long period grating based biosensor for the detection of Escherichia coli bacteria. *Biosens. Bioelectron.* **2012**, *35*, 308–312. [[CrossRef](#)]
46. Chiniforooshan, Y.; Celebanska, A.; Janik, M.; Mikulic, P.; Haddad, F.; Perreault, J.; Bock, W.J. Double resonance long period fiber grating for detection of *E. coli* in trace concentration by choosing a proper bacteriophage. In Proceedings of the 25th International Conference on Optical Fiber Sensors, Jeju, Korea, 24–28 April 2017; Volume 10323.
47. Rossmann, M.G.; Rao, V.B. Advances in Experimental Medicine and Biology. In *Viral Molecular Machines*; Springer: Berlin/Heidelberg, Germany, 2012.
48. Singh, A.; Poshtiban, S.; Evoy, S. Recent Advances in Bacteriophage Based Biosensors for Food-Borne Pathogen Detection. *Sensors* **2013**, *13*, 1763–1786. [[CrossRef](#)] [[PubMed](#)]
49. Brzozowska, E.; Koba, M.; Smietana, M.; Górska, S.; Janik, M.; Gamian, A.; Bock, W.J. Label-free Gram-negative bacteria detection using bacteriophage-adhesin-coated long-period gratings. *Biomed. Opt. Express* **2016**, *7*, 829–840. [[CrossRef](#)]
50. Koba, M.; Smietana, M.; Brzozowska, E.; Górska, S.; Janik, M.; Mikulic, P.; Cusano, A.; Bock, W.J. Bacteriophage Adhesin-Coated Long-Period Grating-Based Sensor: Bacteria Detection Specificity. *J. Lightwave Technol.* **2016**, *34*, 4531–4536. [[CrossRef](#)]

51. Piestrzyńska, M.; Dominik, M.; Kosiel, K.; Janczuk-Richter, M.; Szot-Karpińska, K.; Brzozowska, E.; Shao, L.; Niedziolka-Jonsson, J.; Bock, W.J.; Smietana, M. Ultrasensitive tantalum oxide nano-coated long-period gratings for detection of various biological targets. *Biosens. Bioelectron.* **2019**, *133*, 8–15. [[CrossRef](#)] [[PubMed](#)]
52. Janeway, C.A.; Travers, P.; Walport, M.; Shlomchik, M. *Immunobiology. The Immune System in Health and Disease*, 5th ed.; Garland Publishing: New York, NY, USA, 2001.
53. Welch, N.G.; Scoble, J.; Muir, B.W.; Pigram, P.J. Orientation and characterization of immobilized antibodies for improved immunoassays. *Biointerphases* **2017**, *12*, 02D301. [[CrossRef](#)] [[PubMed](#)]
54. Suo, Z.; Yang, X.; Avci, R.; Deliorman, M.; Rugheimer, P.; Pascual, D.W.; Idzerda, Y. Antibody Selection for Immobilizing Living Bacteria. *Anal. Chem.* **2009**, *81*, 7571–7578. [[CrossRef](#)] [[PubMed](#)]
55. Bandara, A.B.; Zuo, Z.; Ramachandran, S.; Ritter, A.; Heflin, J.; Inzana, T.J. Detection of methicillin-resistant staphylococci by biosensor assay consisting of nanoscale films on optical fiber long-period gratings. *Biosens. Bioelectron.* **2015**, *70*, 433–440. [[CrossRef](#)]
56. Yang, F.; Chang, T.-L.; Liu, T.; Wu, D.; Du, H.; Liang, J.F.; Tian, F.; Yan, F. Label-free detection of *Staphylococcus aureus* bacteria using long-period fiber gratings with functional polyelectrolyte coatings. *Biosens. Bioelectron.* **2019**, *133*, 147–153. [[CrossRef](#)]
57. Kaushik, S.; Tiwari, U.; Nilima, N.; Prashar, S.; Das, B.; Sinha, R.K. Label-free detection of *Escherichia coli* bacteria by cascaded chirped long period gratings immunosensor. *Rev. Sci. Instrum.* **2019**, *90*, 1–9. [[CrossRef](#)]
58. Celebanska, A.; Chiniforooshan, Y.; Janik, M.; Mikulic, P.; Sellamuthu, B.; Perreault, J.; Bock, W.J. Bioinspired Carbohydrate-Decorated Long-Period Fiber Grating for Label-Free Bacteria Detection. *IEEE Sens. J.* **2019**, *19*, 11965–11971. [[CrossRef](#)]
59. Brzozowska, E.; Smietana, M.; Koba, M.; Górska, S.; Pawlik, K.; Gamian, A.; Bock, W.J. Recognition of bacterial lipopolysaccharide using bacteriophage-adhesin-coated long-period gratings. *Biosens. Bioelectron.* **2015**, *67*, 93–99. [[CrossRef](#)] [[PubMed](#)]
60. Koba, M.; Smietana, M.; Brzozowska, E.; Górska, S.; Mikulic, P.; Bock, W.J. Reusable Bacteriophage Adhesin-Coated Long-Period Grating Sensor for Bacterial Lipopolysaccharide Recognition. *J. Lightwave Technol.* **2014**, *33*, 2518–2523. [[CrossRef](#)]
61. Tombelli, S.; Minunni, M.; Mascini, M. Analytical applications of aptamers. *Biosens. Bioelectron.* **2005**, *20*, 2424–2434. [[CrossRef](#)]
62. Song, S.; Wang, L.; Li, J.; Fan, C.; Zhao, J. Aptamer-based biosensors. *TrAC Trends Anal. Chem.* **2008**, *27*, 108–117. [[CrossRef](#)]
63. Queirós, R.; Gouveia, C.; Fernandes, J.R.; Jorge, P.A. Evanescent wave DNA-aptamer biosensor based on long period gratings for the specific recognition of *E. coli* outer membrane proteins. *Biosens. Bioelectron.* **2014**, *62*, 227–233. [[CrossRef](#)]
64. Janik, M.; Koba, M.; Celebańska, A.; Bock, W.J.; Smietana, M. Live, *E. coli* bacteria label-free sensing using a microcavity in-line Mach-Zehnder interferometer. *Sci. Rep.* **2018**, *8*, 17176. [[CrossRef](#)] [[PubMed](#)]
65. Kaushik, S.; Tiwari, U.K.; Pal, S.S.; Sinha, R.K. Rapid detection of *Escherichia coli* using fiber optic surface plasmon resonance immunosensor based on biofunctionalized Molybdenum disulfide (MoS₂) nanosheets. *Biosens. Bioelectron.* **2019**, *126*, 501–509. [[CrossRef](#)]
66. Dandapat, K.; Tripathi, S.M.; Chinifooroshan, Y.; Bock, W.J.; Mikulic, P. Compact and cost-effective temperature-insensitive bio-sensor based on long-period fiber gratings for accurate detection of *E coli* bacteria in water. *Opt. Lett.* **2016**, *41*, 4198. [[CrossRef](#)]
67. Tok, J.B.-H. RNA aptamers that specifically bind to a 16S ribosomal RNA decoding region construct. *Nucleic Acids Res.* **2000**, *28*, 2902–2910. [[CrossRef](#)]



© 2020 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).