

Article

Corrosion Resistance and Surface Bioactivity of Ti6Al4V Alloy after Finish Turning under Ecological Cutting Conditions

Kamil Leksycki ^{1,*}, Agnieszka Kaczmarek-Pawelska ¹, Kamil Ochał ^{2,3}, Andrzej Gradzik ^{2,3},
Danil Yurievich Pimenov ^{4,*}, Khaled Giasin ⁵, Daniel Chuchala ⁶ and Szymon Wojciechowski ^{7,*}

¹ Institute of Mechanical Engineering, University of Zielona Gora, 4 Prof. Z. Szafrana Street, 65-516 Zielona Gora, Poland; a.kaczmarek@iim.uz.zgora.pl

² Department of Materials Science, The Faculty of Mechanical Engineering and Aeronautics, Rzeszow University of Technology, Powstancow Warszawy 12, 35-959 Rzeszow, Poland; kochal@prz.edu.pl (K.O.); andrzej_gradzik@prz.edu.pl (A.G.)

³ Research and Development Laboratory for Aerospace Materials, Rzeszow University of Technology, Zwirki i Wigury 4, 35-036 Rzeszow, Poland

⁴ Department of Automated Mechanical Engineering, South Ural State University, Lenin Prosp. 76, 454080 Chelyabinsk, Russia

⁵ School of Mechanical and Design Engineering, University of Portsmouth, Portsmouth PO1 3DJ, UK; khaled.giasin@port.ac.uk

⁶ Faculty of Mechanical Engineering and Ship Technology, Gdansk University of Technology, 11/12 Gabriela Narutowicza Street, 80-233 Gdansk, Poland; daniel.chuchala@pg.edu.pl

⁷ Faculty of Mechanical Engineering, Poznan University of Technology, 60-965 Poznan, Poland

* Correspondence: k.leksycki@iim.uz.zgora.pl (K.L.); danil_u@rambler.ru (D.Y.P.); sjwojciechowski@o2.pl (S.W.)



Citation: Leksycki, K.; Kaczmarek-Pawelska, A.; Ochał, K.; Gradzik, A.; Pimenov, D.Y.; Giasin, K.; Chuchala, D.; Wojciechowski, S. Corrosion Resistance and Surface Bioactivity of Ti6Al4V Alloy after Finish Turning under Ecological Cutting Conditions. *Materials* **2021**, *14*, 6917. <https://doi.org/10.3390/ma14226917>

Academic Editor: Fuhui Wang

Received: 22 October 2021

Accepted: 11 November 2021

Published: 16 November 2021

Publisher's Note: MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Copyright: © 2021 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Abstract: The influence of cooling conditions and surface topography after finish turning of Ti6Al4V titanium alloy on corrosion resistance and surface bioactivity was analyzed. The samples were machined under dry and minimum quantity lubrication (MQL) conditions to obtain different surface roughness. The surface topographies of the processed samples were assessed and measured using an optical profilometer. The produced samples were subjected to electrochemical impedance spectroscopy (EIS) and corrosion potential tests (E_{corr}) in the presence of simulated body fluid (SBF). The surface bioactivity of the samples was assessed on the basis of images from scanning electron microscopy (SEM) and energy-dispersive spectroscopy (EDS) analysis. The inspection of the surfaces of samples after turning under dry and MQL conditions revealed unevenly distributed precipitation of hydroxyapatite compounds (Ca/P) with a molar ratio in the range of 1.73–1.97. Regardless of the cutting conditions and surface roughness, the highest values of E_{corr} ~0 mV were recorded on day 7 of immersion in the SBF solution. The impedance characteristics showed that, compared to the MQL conditions, surfaces machined under dry conditions were characterized by greater resistance and the presence of a passive layer on the processed surface. The main novelty of the paper is the study of the effect of ecological machining conditions, namely, dry and MQL cutting on the corrosion resistance and surface bioactivity of Ti6Al4V titanium alloy after finish turning. The obtained research results have practical significance. They can be used by engineers during the development of technological processes for medical devices made of Ti6Al4V alloy to obtain favorable functional properties of these devices.

Keywords: Ti6Al4V alloy; finish turning; surface topography; cooling conditions; corrosion resistance; surface bioactivity; simulated body fluid (SBF); minimum quantity lubrication (MQL)

1. Introduction

An important aspect that proves the usability and longevity of metals in the human body is their corrosion resistance [1]. As metal implants are surrounded by aggressive body fluids constituting the internal environment of the human body, their corrosion cannot

be avoided [2]. Ti6Al4V titanium alloy is very commonly used in medical devices and applications [3,4]. In the environment of the human body, it exhibits neutral behavior and high corrosion resistance, making it a biocompatible material [5]. Titanium and its alloys belong to the group of materials with high corrosion resistance, resulting from their ability to passivate, and, in environments containing high concentrations of chloride or sulfide ions, they undergo pitting corrosion [6–9]. An important aspect of implant manufacturing is the quality of the implant surface. For orthopedic implants, it is important to achieve a surface with a designated roughness due to the need for a durable biomaterial–tissue interface. For implants and surgical instruments that are in direct contact with blood, it is important to achieve the lowest possible surface roughness. Producing surfaces with assumed roughness maintains mechanical stability and allows bone tissue to grow into and adhere to the implant surface. Higher implant surface roughness results in more favorable cell adhesion and a more accurate bone–implant fusion [10]. However, higher surface roughness contributes to the development of the implant corrosion process. This is related to the characteristic post-treatment changes present on the surface, which contribute to the formation of corrosion foci [11]. The surface roughness of titanium alloys for both medical and other industries depends on the type and conditions of the treatment it is subjected to [12]. Although titanium alloy Ti6Al4V shows very good mechanical properties and is a widely known material used in industry, it belongs to the group of materials that are difficult to process [13,14]. Due to its properties, high temperature is generated within the cutting zone during machining and adversely affects the quality of the processed surface [15,16].

Previous studies have analyzed the influence of cutting conditions and surface roughness on the corrosion resistance of medical-grade materials in various solutions. Bertolini et al. [17] investigated the corrosion resistance of Ti6Al4V alloy after turning under dry and cryogenic conditions. Cutting speeds (v_c) of 50 and 80 m/min, feed rates (f) of 0.1 and 0.2 mm/rev, and a depth of cut (a_p) of 0.25 mm were applied. The corrosion tests were performed in 0.9% NaCl solution. It was shown that cryogenic processing effectively improves the corrosion and fretting properties of the examined material. Bertolini et al. [18] studied the effect of feed rate and dry and cryogenic conditions on the corrosion behavior of AZ31 alloy. A feed rate within the 0.01–0.1 mm/rev range, cutting speed of 100 m/min, and depth of cut of 0.25 mm were applied. The tests were performed in simulated body fluid (SBF). The feed rate affected the surface roughness and corrosion behavior of AZ31 alloy. The corrosion resistance of AZ31 magnesium alloy was more favorable using a lower feed rate and cryogenic conditions. Pu et al. [19] investigated the corrosion resistance of magnesium alloy AZ31B-O following dry and cryogenic cutting conditions. A tool edge radius of 30 and 70 μm , feed rate of 0.1 mm/rev, and cutting speed of 100 m/min were used. The samples were tested in SBF and 5% NaCl solution. Compared to dry machining, cryogenic ensure better surface integrity and corrosion resistance. Bruschi et al. [20] investigated the corrosion behavior of 316L stainless steel to increase its resistance to sterilization cycles. The processing was carried out under low-temperature conditions, and the results were compared with the dry and wet conditions. The following parameters were applied: $v_c = 200$ m/min, $f = 0.2$ mm/rev, and $a_p = 0.3$ mm. The tests were performed in a commercial IGENAL-N sterilization solution. Corrosion resistance improved with low-temperature conditions and deteriorated after dry conditions. Bertolini et al. [21] analyzed the corrosion behavior of stainless steel 316L following low-temperature and wet conditions. The following parameters were applied: $v_c = 200$ m/min, $f = 0.2$ mm/rev, and $a_p = 0.3$ mm. The tests were performed in a commercial IGENAL-N sterilization solution. Low-temperature conditions ensure a hardened and more elastic surface layer, which, compared to wet conditions, increases the general and local corrosion resistance of the tested steel.

Zhang and Liu [22] evaluated the surface topography of Ti6Al4V alloy after finish turning under dry conditions. The values of R_a and R_z parameters increased, and the shapes of surface roughness profiles changed with increasing feed rate. Sartori et al. [23] investigated the surface integrity of Ti6Al4V alloy after cryogenic and dry cutting conditions. Surface topography deteriorated with cryogenic conditions, but fewer surface defects were observed. Mia et al. [24] investigated the surface roughness of Ti6Al4V alloy after turning under cryogenic liquid nitrogen (LN2) conditions. LN2 dual jets improve the surface quality. The following cutting parameters were recommended: $v_c = 140$ m/min and $f = 0.16$ mm/rev. Deiab et al. [25] assessed the effect of dry, cooled air, flood, cryogenic, minimum quantity lubrication (MQL), and minimum quantity cooled lubrication (MQCL) conditions on the surface roughness of Ti6Al4V alloy. At a higher machining speed and feed rate in dry and MQL conditions, the lowest R_a parameter values were obtained. Sun et al. [26] analyzed the surface roughness of Ti-5553 alloy under cryogenic conditions, and the results were compared to wet and MQL conditions. The smallest values of the R_a parameter were obtained from MQL. Sun et al. [27] investigated the surface integrity after turning titanium alloy TB6 under dry conditions. A decrease in the R_a parameter value was observed with decreasing feed rate. Fewer defects on the surface were observed with increasing cutting speed. Gupta et al. [28] analyzed the effect of Ranque-Hilsch vortex fluid-assisted minimum quantity cutting fluids (RHVT-MQCF) and MQL cutting conditions on surface roughness after turning pure titanium. Favorable results were obtained for the machining conditions with RHVT-MQCF.

There are few studies in the open literature which looked into the effect of surface topography and cutting conditions on the corrosion resistance after finishing turning of titanium alloys. Yet, the processing conditions affect both the surface topography [29] and the corrosion resistance [30]. The existing articles do not consider the influence of modern cooling conditions, i.e., near-dry cutting, which improves the surface quality [31] while offering numerous economic and environmental benefits [32]. In addition, there is no information on the effect of surface topography and cooling conditions on surface bioactivity after the turning process of titanium alloys, which can be considered a novelty in this research work. As studies have shown, surface topography affects surface bioactivity. Ravelingien et al. [33] investigated the effect of surface topography on the surface bioactivity of Ti6Al4V alloy after alkali treatment. The titanium plates had different surface roughness ($R_a = 0.13$ μm , 0.56 μm , 0.83 μm , and 3.63 μm) and were prepared by Al_2O_3 grit-blasting. For samples with $R_a = 0.56$ μm , complete hydroxyapatite (HA) coating was obtained after 7 days of storage in SBF. Constantinescu et al. [34] studied the effect of surface roughness of Ti6Al4V alloy on the shape, size, and distribution of HA. The samples were sandblasted with Al_2O_3 particles of different sizes. It was found that surface roughness affects the morphology and distribution of HA. Demirci et al. [35] analyzed the effect of surface modifications of Ti6Al4V alloy on the ability to form apatite. Samples were produced using the additive method (AM) using varying laser powers. The samples were stored in SBF for 2 weeks. Depending on the production laser power, surface roughness and topography, as well as wettability and microstructure, affected the formation of apatite on the surface.

A high surface bioactivity of titanium alloys, shaped by various techniques to produce orthopedic implants, is achieved by anodizing the surface in solutions containing phosphate ions or by forming layers with morphology favoring the surface bioactivity [36]. Increasing the surface bioactivity of titanium implants can also be achieved by covering their surface with a layer containing calcium and phosphorus hydroxyapatite. HA-coated titanium alloys exhibit higher surface bioactivity than uncoated ones [37]. HA coating aids implant fixation and increases implant life. The dense layer of HA coating on the implant surface has a beneficial effect on biointegration, improves the stability of the implant and bone tissue junction, and minimizes the release of metal ions in the human body [38]. Hydroxyapatite is a ceramic material that is a natural bone



component. In the assessment of the in vitro surface bioactivity of titanium alloys, it was verified, inter alia, whether hydroxyapatite is released on a given surface of the alloy following processing [39]. HA is like a bioceramic that has very good biological properties facilitating bone repair and reconstruction [40]. The surface roughness of titanium alloys suitable for the growth of osteoblasts is 0.2–0.7 μm , and, in the literature, many reports can be found on the protective and bioactive layers on titanium alloys and their corrosion resistance [41–44].

Summarizing the literature review, it was found that, in the research works to date, there is a lack of information on the effect of finish turning of Ti6Al4V titanium alloy under ecological conditions on the corrosion resistance and surface bioactivity in SBF solution, whose chemical composition is most similar to human blood plasma. In the research, the simultaneous effect of feed rate and surface roughness on these performance properties of Ti6Al4V titanium alloy has also basically not been analyzed. This can be considered a research gap. Therefore, the current study aimed to analyze the effect of dry and MQL conditions and surface topography after finish turning of Ti6Al4V titanium alloy on corrosion resistance and surface bioactivity in SBF solution. This research is of practical importance and can be used as a step in the development of technological processes in the production of medical devices, to ensure favorable operating properties of these devices.

2. Materials and Methods

2.1. Workpiece Details and Preparation

Ti6Al4V alloy, with the chemical composition and mechanical properties compliant with ISO 5832-3:2016 [45] standard, was examined (Table 1). Ti6Al4V alloy is characterized by high strength-to-weight ratios and has high corrosion resistance; thus, it is used in many industries including medical, automotive, aerospace, and marine [46].

Table 1. Chemical composition and mechanical properties of Ti6Al4V titanium alloy, According to ISO 5832-3:2016.

Chemical Composition (%)							
O	V	Al	Fe	H	C	N	Ti
<0.20	3.5	5.5	<0.30	<0.0015	<0.08	<0.05	rest
Mechanical Properties							
Modulus of Elastic (MPa)		Tensile Strength (MPa)		Yield Strength (MPa)		Fatigue Strength (MPa)	
110–114		960–970		850–900		620–725	

The machining was performed on a CNC lathe CKE6136i (Dalian Machine Tool Group Corporation, Dalian, China) using a turning with a CoroTurn SDJCR 2020K 11 (Sandvik Coromant, Sandviken, Sweden) tool holder and a CoroTurn DCMX 11 T3 04-WM 1115 (Sandvik Coromant, Sandviken, Sweden) insert. The geometry of the cutting insert was as follows: $\kappa_r = 93^\circ$, $\gamma_o = 18^\circ$, $\alpha_o = 7^\circ$, $r_\epsilon = 0.4$ mm. The material used for the cutting insert was cemented carbide GC 1115 with (Ti, Al)N + (Al, Cr)₂O₃ coating deposited by Physical Vapour Deposition (PVD). The inserts were changed after each test, so each sample was machined with a new and sharp edge. The samples were machined under dry and MQL conditions because today's manufacturing industry is very much focused on environmental protection [47] and a reduction in the consumption of machining fluids [48]. Furthermore, MQL has been successfully applied to improve the machinability of difficult-to-cut materials [49,50]. In the MQL method, we used ECOCUT MIKRO 20 E oil, which was produced by mixing with air. The oil mist was produced using the Lenox 1LN Mikronizer (Lenox, East Longmeadow, MA, USA). The preferred oil mist formation conditions suggested by Maruda et al. were applied [51]: oil flow 39.4 mL/h, airflow 5.8 L/m, pressure 0.48 MPa, and 0.2 m distance of the nozzle from the point of contact between the material and the cutting edge. The samples

with minimum ($R_{a\min} = 0.29\text{--}0.37 \mu\text{m}$) and maximum ($R_{a\max} = 1.62\text{--}2.22 \mu\text{m}$) surface roughness obtained after finish turning were examined. Surface roughness affects the corrosion of metals and, as commonly known, depends on the feed rate (f). For samples with $R_{a\min}$, a lower feed rate was used for MQL conditions ($f = 0.1 \text{ mm/rev}$) to obtain surface roughness parameters similar to dry machining ($f = 0.14 \text{ mm/rev}$). Regardless of the machining conditions, the feed rate for the samples with $R_{a\max}$ was 0.35 mm/rev . Due to the small diameter of the sample, a constant cutting speed (v_c) of 80 m/min was set, as well as a constant cutting depth (a_p) of 0.5 mm , typical for the finish cutting. Both the length and the diameter of the samples were 15 mm .

2.2. Measurement of Surface Roughness Metrics

The processed surface was examined using a Sensofar S Neox 3D optical profilometer (Sensofar Group, Barcelona, Spain), and the measurement results were analyzed using Mountains Maps Premium 7.4 software (Digital Surf, Besançon, France). The measurement area used was $1.30 \times 1.75 \text{ mm}$. Selected parameters of the surface roughness amplitude were analyzed in accordance with the ISO 4287:1999 [52] standard R_a (arithmetic mean deviation of the roughness profile), R_z (maximum height of the roughness profile), R_p (maximum peak height of the roughness profile), R_v (maximum valley depth of the roughness profile), R_{sk} (skewness of the roughness profile), R_{ku} (kurtosis of the roughness profile), roughness profiles, and 2D topographies of the processed surface. In industry, parameters R_a and R_z are widely used to evaluate product quality. The parameters R_{sk} and R_{ku} can be complementary to the parameters R_a and R_z and allow the evaluation of the functional properties of the surface [53,54]. The tests were repeated three times and the standard deviation did not exceed 5%.

Before corrosion resistance and surface bioactivity testing, the samples were washed in an IS-1 (Intersonic, Olsztyn, Poland) ultrasonic washer in distilled water to remove machining residues, such as machined material or oil that sticks to the surface. Samples of Ti6Al4V titanium alloy were subjected to electrochemical tests using the Atlas 0531 Electrochemical Unit and Impedance Analyzer potentiostat/galvanostat (Atlas-Sollich, Rebiechowo, Poland). The chemical composition of the simulated body fluid can be found in [55]. The SBF used in this study had a temperature of $37 \pm 1 \text{ }^\circ\text{C}$, and a pH in the range of 7.2–7.4. An area of 0.25 cm^2 was marked on each sample, which was then exposed to the electrolyte. The remainder of the sample was covered with a layer of paint. The tests were repeated three times and the standard deviation did not exceed 3%.

The samples were subjected to electrochemical impedance spectroscopy (EIS) tests, preceded by a 1 min measurement of the corrosion potential (E_{corr}). EIS studies were performed according to ISO 16773-2:2016 [56] and ISO 16773-3:2016 [57] standards. The tests were carried out in a three-electrode system, in which the reference electrode was silver chloride, with a platinum plate as the counter electrode. Changes in electrochemical properties of the samples were recorded after 1 h and then after 1, 7, 14, 28, 46, and 72 days of immersion in SBF solution. EIS spectra were recorded within the $10^{-5}\text{--}0.1 \text{ Hz}$ frequency range. The results obtained were analyzed, and Bode and Nyquist charts were plotted. The sample surfaces were assessed on the basis of SEM images and EDS analysis obtained using a JSM-5600LV scanning microscope (JEOL, Tokyo, Japan) with an EDS 2000 X-ray microanalyzer and an AVT-HORN camera (AVT, Aachen, Germany). To neutralize the influence of Ti, Al, and V on the effectiveness of registration of the presence of hydroxyapatite compounds on the surfaces of the tested titanium alloy samples, analyzed by EDS, these elements were excluded from the list of analyzed elements. A similar approach was put forward by Feldshtein et al. [58].

The experimental research plan is shown in Figure 1.




Figure 1. Experimental research plan.

3. Results and Discussion

Surface roughness parameters were obtained for samples with minimum and maximum R_a for Ti6Al4V alloy after finish turning under the ecological cutting conditions shown in Table 2.

Table 2. Surface Roughness Parameters for Samples with R_{amin} and R_{amax} Obtained after Finish Turning.

Cutting Conditions	Surface Roughness Range	R_a (μm)	R_z (μm)	R_p (μm)	R_v (μm)	R_{sk}	R_{ku}
Dry	R_{amin}	0.29	2.18	1.16	1.01	-0.208	2.68
MQL		0.37	2.33	1.27	1.07	0.135	2.49
Dry	R_{amax}	2.22	11.8	7.23	4.62	0.287	2.58
MQL		1.62	8.80	5.43	3.37	0.311	2.64

Garcia-Martinez et al. [59] showed that machining under MQL conditions reduces surface roughness and improves surface integrity compared to dry machining. For samples with R_{amin} after dry turning, a reduction in surface roughness parameters R_a , R_z , R_p , R_v , and R_{sk} in the range of 6% to 165% was obtained compared to MQL conditions. On the other hand, for samples with R_{amax} machined under MQL conditions, a reduction in R_a , R_z , R_p , and R_v parameters was obtained in the range of 25% to 27%.

Figure 2 shows the surface roughness profiles of samples with R_{amin} and R_{amax} parameters obtained after finish turning under dry and MQL conditions.

Within the range of minimum parameter values, $R_a = 0.29\text{--}0.37 \mu\text{m}$ and $R_z = 2.18\text{--}2.33 \mu\text{m}$, regardless of the cooling conditions, flattened peaks were observed on the surface roughness profiles attesting to the plastic deformation of the surface layers characteristic of the processed implant surface [60]. However, within the range of maximum $R_a = 1.62\text{--}2.22 \mu\text{m}$ and $R_z = 8.80\text{--}11.08 \mu\text{m}$ parameter values, there were visible feed marks. For example, Figure 2 shows a cutout of the 2D surface topography along with the surface roughness profile for a R_{amax} sample after dry cutting.




Figure 2. Surface roughness profiles of Ti6Al4V alloy after finish turning.

The correlation of the R_{sk} and R_{ku} parameters yields an R_{sk} – R_{ku} topological map, which enables an evaluation of the surface irregularity distribution. The parameters R_{sk} and R_{ku} are responsive to high peaks and deep pits located on the machined surface [61]. The R_{sk} – R_{ku} topological map for Ti6Al4V alloy after finish turning is shown in Figure 3. Compared to the surface area of samples with $R_{a\min}$, an increase in R_{sk} parameters was obtained for samples with $R_{a\max}$. Very high peaks and very deep pits were seen on surfaces with $R_{a\max}$, whereas high peaks and deep pits were seen on surfaces with $R_{a\min}$ machined under MQL conditions. In turn, flattened peaks were seen on the surface with $R_{a\min}$ under dry machined conditions. Regardless of the cooling conditions for the samples with $R_{a\min}$ and $R_{a\max}$, the surfaces were characterized by regular shapes, as evidenced by the parameter value $R_{ku} < 3$.




Figure 3. R_{ku} – R_{sk} topological map for Ti6Al4V alloy after finish turning.

Figure 4 shows 2D surface images of the samples with R_{amin} and R_{amax} after finishing turning. For R_{amin} , on the surface processed under dry and MQL conditions, scratches were observed, which were caused by the unfavorable shape of the chip entangling the workpiece [62,63] and irregularly spaced stickers, resulting from the adhesive bonds of the chips to the processed surface [64]. For R_{amax} , on the surface processed in the tested cooling conditions, clear traces of feed rate, which are typical of turning, were found. In addition, scratches occurred under dry cutting conditions and stickers with MQL. A summary of the surface topography results of samples with R_{amin} and R_{amax} of Ti6Al4V alloy after turning under dry and MQL conditions is presented in Table 3.




Figure 4. The 2D topographies of Ti6Al4V titanium alloy surface after finish turning.

Table 3. Summary of Surface Topography Results of Samples with R_{amin} and R_{amax} of Ti6Al4V Alloy after Finish Turning under Ecological Cutting Conditions.

Cutting Conditions	Surface Roughness Range	Surface Roughness Profiles	Surface Roughness Parameters	R_{ku} – R_{sk} Topological Maps	2D Surface Images
Dry	R_{amin}	Flattened peaks	Dry compared to MQL: decrease R_a , R_z , R_p , R_v , and R_{sk} of 6–165%	Flattened peaks, regular shapes	Scratch, stickers
MQL					
Dry	R_{amax}	Feed marks	MQL compared to dry: decrease R_a , R_z , R_p , and R_v of 25–27%	Very high peaks and very deep pits, regular shapes	Feed marks, scratch
MQL					

The results of the impedance tests showed that the cutting conditions and surface topography after finish turning of Ti6Al4V alloy affected the bioactivity of the surface and the release of hydroxyapatite and changed the electrochemical characteristics of the layer. SEM images and EDS analyses of the surfaces after finish turning after dry and MQL conditions after 7 days of storage in SBF solution are shown in Figure 5.




Figure 5. SEM pictures and EDS analysis results for Ti6Al4V alloy with $R_{a\min}$ and $R_{a\max}$ under dry and MQL conditions after 7 days of storage in SBF.

Spherical precipitates unevenly covering the surface of the samples were observed for samples after turning under dry and MQL conditions. Analysis of the content of calcium and phosphorus confirmed the presence of hydroxyapatite, in which the Ca/P molar ratio was 1.73–1.78 for samples with $R_{a\min}$ and 1.85–1.97 for samples with $R_{a\max}$. It should be emphasized that the secretions of hydroxyapatite were more irregular on samples after turning with MQL conditions than on samples subjected to dry machining; furthermore, their stoichiometric composition was closest to the composition of hydroxyapatite, which is a bone component (Ca/P = 1.67) [36]. Changes in the corrosion potential of samples made of Ti6Al4V alloy with $R_{a\min}$ and $R_{a\max}$ after turning under dry and MQL conditions stored in SBF solution are shown in Figure 6a.

For samples with $R_{a\min}$ and $R_{a\max}$ after turning under dry and MQL conditions, over time after immersion in SBF, the values of the corrosion potentials decreased and then significantly increased, to reach the values of $E_{\text{corr}} \sim 0$ mV. Then, they decreased to $E_{\text{corr}} \sim -120$ to -140 mV, which continued until the 72nd day. The increase in the value of corrosion potentials on the seventh day from the moment of immersion in SBF proves the existence of positive ion adsorption processes from the solution, forming a double electrical layer at the alloy surface, from which hydroxyapatite is released [65]. The results obtained in the first hours/days after immersion in the solution are very important, as they may indicate the body's reaction to the metal placed in it as an implant [66].

The results of impedance tests for samples made of Ti6Al4V alloy with $R_{a\min}$ and $R_{a\max}$ after turning with dry and MQL conditions after 7 days of storage in SBF solution are presented in the form of Bode and Nyquist diagrams in Figures 6 and 7. The impedance characteristics (Figures 6 and 7) indicate that, regardless of the surface roughness of samples, compared to MQL, under dry conditions, surfaces were characterized by a greater resistance and an almost capacitive response, illustrated by a phase angle close to -80° recorded in a wide frequency range by control systems including in the range of 10^{-1} – 10^3 Hz, indicating the presence of a passive layer on the processed surface. The test results for the anodized titanium alloy Ti4Al4V reported in the literature show a similar characteristic [12,37,39]. In the Nyquist illustration, samples after turning under MQL conditions were characterized

by one loop (Figure 7b); when compared to samples with $R_{a\max}$, higher resistance values were recorded for samples with $R_{a\min}$. It is reported in the literature that, in the case of polished samples, larger diameters of the loops are observed, which are evidence of a higher resistance [12]. For samples with $R_{a\max}$, the lowest resistance occurred 7 days after immersion in SBF solution. Then, after 28 days, the resistance value increased and remained at a similar level for 72 days, suggesting that no hydroxyapatite was recorded on the surfaces of the samples during this period. The obtained data were then filtered. The equivalent circuit Rs (P1, R1(P2, R2)) provided the best fit to the test data, where Rs is the electrolyte resistance, R1 and CPE1 are the hydroxyapatite layer resistivity and metal capacity, and R2 and CPE2 are the charge transfer resistance and double-layer capacity (Figure 7e). A summary of the corrosion resistance and surface bioactivity test results of samples with $R_{a\min}$ and $R_{a\max}$ of Ti6Al4V alloy after turning under dry and MQL conditions is presented in Table 4.




Figure 6. The results of electrochemical tests of Ti6Al4V alloy samples after turning under dry and MQL conditions with $R_{a\min}$ and $R_{a\max}$ stored in SBF: changes in corrosion potential for 72 days (a) the value of the phase angle at the frequency of 0.1 Hz after 7 days (b) the Bode phase plots after 7, 28, and 72 days (c,d) and the Bode plots after 7, 28, and 72 days (e,f).




Figure 7. Nyquist charts for samples of Ti6Al4V alloy after turning with $R_{a\min}$ and $R_{a\max}$ after 7, 28, and 72 days of storage in SBF, under dry (a) and MQL (b) conditions; comparison of impedance characteristics in Nyquist representation (c) and Bode phase plots (d) with equivalent circuit illustrating the nature of samples (e).

Table 4. Summary of Corrosion Resistance and Surface Bioactivity Results of Samples with $R_{a\min}$ and $R_{a\max}$ of Ti6Al4V Alloy after Turning under Ecological Cutting Conditions.

Cutting Conditions	Surface Roughness Range	Precipitation of Hydroxyapatite, after 7 Days	Stoichiometric Composition (Ca/P), after 7 Days	Highest Values of E_{corr}	Impedance Characteristics
Dry	$R_{a\min}$	Irregular, spherical	1.78	~0 mV, after 7 days	Higher resistance, an almost capacitive response, presence of a passive layer
MQL		More irregular, spherical	1.73		Low resistance, lack of presence of a passive layer
Dry	$R_{a\max}$	Irregular, spherical	1.97		High resistance, an almost capacitive response, presence of a passive layer
MQL		More irregular, spherical	1.85		Low resistance, lack of presence of a passive layer

4. Conclusions

The aim and novelty of this work was to analyze the effect of ecological dry and MQL machining conditions and surface topography of Ti6Al4V titanium alloy after finish turning on corrosion resistance and surface bioactivity. The surface topographies of the processed samples were assessed and measured using a Sensofar S Neox 3D optical profilometer. Samples with different surface roughness were subjected to electrochemical impedance spectroscopy and corrosion potential tests in the presence of simulated body fluid. The following was established:

1. For samples with $R_{a\min}$ after dry turning, a reduction in surface roughness parameters of 6% to 165% was obtained compared to MQL conditions. On the other hand, for samples with $R_{a\max}$ machined under MQL conditions, a reduction was obtained of 25% to 27%.
2. Regardless of the cutting conditions within the minimum range $R_a = 0.25\text{--}0.37\ \mu\text{m}$, flattened peaks were observed on the surface roughness profiles, and, within the range of the maximum $R_a = 1.62\text{--}2.22\ \mu\text{m}$, there were visible feed marks.
3. $R_{sk}\text{--}R_{ku}$ topological map showed very high peaks and very deep pits on surfaces with $R_{a\max}$, as well as high peaks and deep pits on surfaces with $R_{a\min}$, machined under MQL conditions. Flattened peaks were seen on the surface with $R_{a\max}$ under dry machined conditions. On the 2D images of the surface with $R_{a\max}$, clear traces of the feed rate were recorded, and, on the surfaces with $R_{a\min}$, scratches and irregularly spaced stickers were observed under the analyzed cutting conditions.
4. After turning under dry and MQL conditions, unevenly distributed precipitates of hydroxyapatite compounds were present on the surfaces of the samples. The Ca/P molar ratio for samples with $R_{a\min}$ was within the range 1.73–1.78, whereas that for samples with $R_{a\max}$ was within the range 1.85–1.97.
5. For the studied cutting conditions and surface roughness, the highest values of E_{corr} ~0 mV were recorded on day 7 of immersion in the SBF solution.
6. The impedance characteristics indicated that, compared to MQL conditions, under dry conditions, surfaces were characterized by a greater resistance and an almost capacitive response, illustrated by a phase angle close to -80° recorded in a wide frequency range by control systems including in the range of $10^{-1}\text{--}10^3$ Hz, indicating the presence of a passive layer on the processed surface.
7. The obtained research results have practical significance. They can be used by engineers during the development of technological processes for medical devices made of Ti6Al4V alloy, to obtain favorable functional properties of these devices, i.e., corrosion resistance and bioactivity of the surface after finish turning. Therefore, a lower surface roughness under dry conditions is recommended to achieve this success.

Author Contributions: Conceptualization, K.L.; methodology, K.L. and A.K.-P.; software, K.L. and A.K.-P.; validation, A.K.-P.; formal analysis, K.L. and A.K.-P.; investigation, K.L. and A.K.-P.; data curation, K.L., A.K.-P., K.O., and A.G.; writing—original draft preparation, K.L. and A.K.-P.; writing—review and editing, K.L., A.K.-P., D.Y.P., K.G., D.C., and S.W.; visualization, K.L., K.O., and A.G.; supervision, K.L. and A.K.-P.; project administration, D.Y.P., K.G., D.C., and S.W. All authors have read and agreed to the published version of the manuscript.

Funding: This work was supported by the Polish Minister of Science and Higher Education (003/RID/2018/19).

Institutional Review Board Statement: Not applicable.

Informed Consent Statement: Not applicable.

Conflicts of Interest: The authors declare no conflict of interest.

Nomenclature

a_p	Depth of cut (mm)
E_{corr}	Corrosion potential (mV)
EDS	Energy dispersive spectroscopy
EIS	Electrochemical impedance spectroscopy
f	Feed rate (mm/rev)
HA	Hydroxyapatite
MQL	Minimum quantity lubrication
R_a	Arithmetic mean deviation of the roughness profile (μm)
R_{amin}	The minimum value of the surface roughness parameter R_a (μm)
R_{amax}	The maximum value of the surface roughness parameter R_a (μm)
R_{ku}	Kurtosis of the roughness profile
R_p	Maximum peak height of the roughness profile (μm)
R_{sk}	Skewness of the roughness profile
R_v	Maximum valley depth of the roughness profile (μm)
R_z	Maximum height of the roughness profile (μm)
SBF	Simulated body fluid
SEM	Scanning electron microscope
v_c	Cutting speed (m/min)

References

- Uo, M.; Watari, F.; Yokoyama, A.; Matsuno, H.; Kawasaki, T. Visualization and detectability of elements rarely contained in soft tissue by X-ray Canning analytical microscopy and elektron-probe micro analysis. *Biomaterials* **2001**, *22*, 1787–1794. [[CrossRef](#)]
- Manam, N.S.; Harun, W.S.W.; Shri, D.N.A.; Ghani, S.A.C.; Kurniawan, T.; Ismail, M.H.; Ibrahim, M.H.I. Study of corrosion in biocompatible metals for implants: A review. *J. Alloy. Compd.* **2017**, *701*, 698–715. [[CrossRef](#)]
- Geetha, M.; Singh, A.K.; Asokamani, R.; Gogia, A.K. Ti based biomaterials, the ultimate choice for orthopaedic implants—A review. *Prog. Mater. Sci.* **2009**, *54*, 397–425. [[CrossRef](#)]
- Pimenov, D.Y.; Mia, M.; Gupta, M.K.; Machado, A.R.; Tomaz, Í.V.; Sarikaya, M.; Wojciechowski, S.; Mikolajczyk, T.; Kaplonek, W. Improvement of machinability of Ti and its alloys using cooling-lubrication techniques: A review and future prospect. *J. Mater. Res. Technol.* **2021**, *11*, 719–753. [[CrossRef](#)]
- Costa, B.C.; Tokuhara, C.K.; Rocha, L.A.; Oliveira, R.C.; Lisboa-Filho, P.N.; Pessoa, J.C. Vanadium ionic species from degradation of Ti-6Al-4V metallic implants: In Vitro cytotoxicity and speciation evaluation. *Mater. Sci. Eng. C* **2019**, *96*, 730–739. [[CrossRef](#)]
- Schenk, R. The Corrosion Properties of Titanium and Titanium Alloys. In *Titanium in Medicine. Engineering Materials*; Springer: Berlin/Heidelberg, Germany, 2001. [[CrossRef](#)]
- Bartmański, M.; Pawłowski, Ł.; Belcarz, A.; Przekora, A.; Ginalska, G.; Strugała, G.; Cieślík, B.; Pałubicka, A.; Zieliński, A. The Chemical and Biological Properties of Nanohydroxyapatite Coatings with Antibacterial Nanometals, Obtained in the Electrophoretic Process on the Ti13Zr13Nb Alloy. *Int. J. Mol. Sci.* **2021**, *22*, 3172. [[CrossRef](#)]
- Pawłowski, Ł.; Bartmański, M.; Mielewczyk-Gryń, A.; Cieślík, B.; Gajowiec, G.; Zieliński, A. Electrophoretically Deposited Chitosan/Eudragit E 100/AgNPs Composite Coatings on Titanium Substrate as a Silver Release System. *Materials* **2021**, *14*, 4533. [[CrossRef](#)] [[PubMed](#)]
- Laboulaisa, J.N.; Mata, J.A.A.; Borrás, V.A.; Muñoz, A.I. Electrochemical characterization and passivation behaviour of new beta-titanium alloys (Ti35Nb10Ta-xFe). *Electrochim. Acta* **2017**, *227*, 410–418. [[CrossRef](#)]

10. Czarnowska, E.; Wierzchoń, T.; Maranda-Niedbala, A.; Karczarewicz, E. Improvement of titanium alloy for biomedical applications by nitriding and carbonitriding processes under glow discharge conditions. *J. Mater. Sci. Mater. Med.* **2000**, *11*, 73–81. [[CrossRef](#)] [[PubMed](#)]
11. Gotz, H.E.; Muller, M.; Emmel, A.; Holzwarth, U.; Erben, R.G.; Stangl, R. Effect of surface finish on the osseointegration of laser-treated titanium alloy implants. *Biomaterials* **2004**, *25*, 4057–4064. [[CrossRef](#)]
12. Chi, G.; Yi, D.; Liu, H. Effect of roughness on electrochemical and pitting corrosion of Ti-6Al-4V alloy in 12 wt. % HCl solution at 35 °C. *J. Mater. Res. Technol.* **2020**, *9*, 1162–1174. [[CrossRef](#)]
13. Childs, T.H.C.; Arrazola, P.J.; Aristimuno, P.; Garay, A.; Sacristan, I. Ti6Al4V metal cutting chip formation experiments and modeling over a wide range of cutting speeds. *J. Mater. Process. Technol.* **2018**, *255*, 898–913. [[CrossRef](#)]
14. Sarikaya, M.; Gupta, M.K.; Tomaz, I.; Pimenov, D.Y.; Kuntoğlu, M.; Khanna, N.; Yıldırım, Ç.V.; Krolczyk, G.M. A state-of-the-art review on tool wear and surface integrity characteristics in machining of superalloys. *CIRP J. Manuf. Sci. Technol.* **2021**, *35*, 624–658. [[CrossRef](#)]
15. Michailidis, N. Variations in the cutting performance of PVD-coated tools in milling Ti6Al4V, explained through temperature-dependent coating properties. *Surf. Coat. Technol.* **2016**, *304*, 325–329. [[CrossRef](#)]
16. Gupta, M.K.; Song, Q.; Liu, Z.; Sarikaya, M.; Mia, M.; Jamil, M.; Singla, A.K.; Bansal, A.; Pimenov, D.Y.; Kuntoğlu, M. Tribological performance based machinability investigations in cryogenic cooling assisted turning of α - β titanium Alloy. *Tribol. Int.* **2021**, *160*, 107032. [[CrossRef](#)]
17. Bertolini, R.; Bruschi, S.; Bordin, A.; Ghiotti, A.; Pezzato, L.; Dabala, M. Fretting Corrosion Behavior of Additive Manufactured and Cryogenic-Machined Ti6Al4V for Biomedical Applications. *Adv. Eng. Mater.* **2017**, *19*, 1500629. [[CrossRef](#)]
18. Bertolini, R.; Bruschi, S.; Ghiotti, A.; Pezzato, L.; Dabala, M. The effect of cooling strategies and machining feed rate on the corrosion behavior and wettability of AZ31 alloy for biomedical applications. *Procedia Cirp* **2017**, *65*, 7–12. [[CrossRef](#)]
19. Pu, Z.; Dillon, O.W.; Puelo, D.A.; Jawahir, I.S. Cryogenic machining and burnishing of magnesium alloys to improve In Vivo corrosion resistance. In *Surface Modification of Magnesium and its Alloys for Biomedical Applications. Volume II: Modification and Coating Techniques Wood head Publishing Series in Biomaterials*; Woodhead Publishing: Sawston, UK, 2015; pp. 103–133. [[CrossRef](#)]
20. Bruschi, S.; Pezzato, L.; Ghiotti, A.; Dabala, M.; Bertolini, R. Effectiveness of using low-temperature coolants in machining to enhance durability of AISI 316L stainless steel for reusable biomedical devices. *J. Manuf. Process.* **2019**, *39*, 295–304. [[CrossRef](#)]
21. Bertolini, R.; Bruschi, S.; Ghiotti, A. Enhancement of corrosion resistance to sterilization stages of a biomedical grade AISI 316L stainless steel by means of low-temperature machining. *Mater. Today* **2019**, *7*, 552–559. [[CrossRef](#)]
22. Zhang, P.; Liu, Z. Modeling and prediction for 3D surface topography in finish turning with conventional and wiper inserts. *Measurement* **2016**, *94*, 37–45. [[CrossRef](#)]
23. Sartori, S.; Bordin, A.; Ghiotti, A.; Bruschi, S. Analysis of the surface integrity in cryogenic turning of Ti6Al4V produced by Direct Melting Laser Sintering. *Procedia Cirp* **2016**, *45*, 123–126. [[CrossRef](#)]
24. Mia, M.; Gupta, M.K.; Lozano, J.A.; Carou, D.; Pimenov, D.Y.; Królczyk, G.; Khan, A.M.; Dhar, N.R. Multi-objective optimization and life cycle assessment of eco-friendly cryogenic N₂ assisted turning of Ti-6Al-4V. *J. Clean. Prod.* **2019**, *210*, 121–133. [[CrossRef](#)]
25. Deiab, I.; Raza, S.W.; Pervaiz, S. Analysis of Lubrication Strategies for Sustainable Machining during Turning of Titanium Ti-6Al-4V Alloy. *Procedia CIRP* **2014**, *17*, 766–771. [[CrossRef](#)]
26. Sun, Y.; Huang, B.; Puleo, D.A.; Jawahir, I.S. Enhanced machinability of Ti-5553 alloy from cryogenic machining: Comparison with MQL and flood-cooled machining and modeling. *Procedia CIRP* **2015**, *31*, 477–482. [[CrossRef](#)]
27. Sun, Y.; Huang, S.; Wang, T.; Chen, W. Research on surface integrity of turning titanium alloy TB6. *Procedia Cirp* **2018**, *71*, 484–489. [[CrossRef](#)]
28. Gupta, M.K.; Song, Q.; Liu, Z.; Pruncu, C.I.; Mia, M.; Singh, G.; Lozano, J.A.; Carou, D.; Khan, A.M.; Jamil, M.; et al. Machining characteristics based life cycle assessment in eco-benign turning of pure titanium alloy. *J. Clean. Prod.* **2020**, *251*, 119598. [[CrossRef](#)]
29. Cai, Y.; Liu, Z.; Shi, Z.; Song, Q.; Wan, Y. Influence of machined surface roughness on thrust performance of micro-nozzle manufactured by micro-milling. *Exp. Therm. Fluid Sci.* **2016**, *77*, 295–305. [[CrossRef](#)]
30. Yang, J.; Wang, X.; Kang, M. Finite element simulation of surface roughness in diamond turning of spherical surfaces. *J. Manuf. Process.* **2018**, *31*, 768–775. [[CrossRef](#)]
31. Park, K.-Y.; Olortegui-Yume, J.; Yoon, M.-C.; Kwon, P. A study on droplets and their distribution for minimum quantity lubrication (MQL). *Int. J. Mach. Tools Manuf.* **2010**, *50*, 824–833. [[CrossRef](#)]
32. Krolczyk, G.M.; Maruda, R.W.; Krolczyk, J.B.; Wojciechowski, S.; Mia, M.; Nieslony, P.; Budzik, G. Ecological trends in machining as a key factor in sustainable production—A review. *J. Clean. Prod.* **2019**, *218*, 601–615. [[CrossRef](#)]
33. Ravelingien, M.; Hervent, A.S.; Mullens, S.; Luyten, J.; Vervaet, C.; Remon, P.J. Influence of surface topography and pore architecture of alkali-treated titanium on In Vitro apatite deposition. *Appl. Surf. Sci.* **2010**, *256*, 3693–3697. [[CrossRef](#)]
34. Demirci, S.; Dalmis, R.; Dikici, T.; Tuncay, M.M.; Kaya, N.; Gulluoglu, A.N. Effect of surface modifications of additively manufactured Ti-6Al-4V alloys on apatite formation ability for biomedical applications. *J. Alloy. Compd.* **2021**, *887*, 161445. [[CrossRef](#)]
35. Constantinescu, F.; Ciocoiu, R.; Trante, O.; Ciuca, I. Hydroxyapatite coating influenced by Ti6Al4V substrate roughness. *Key Eng. Mater.* **2015**, *638*, 62–66. [[CrossRef](#)]
36. Kaczmarek-Pawelska, A.; Krasicka-Cydzik, E. Functionalized nanotubular oxide layer on Ti6Al4V as a regulator and biosensor of bone tissue remodelling. *Arch. Mater. Sci. Eng.* **2015**, *73*, 53–61.

37. Yigit, O.; Dikici, B.; Senocak, T.C.; Ozdemir, N. One-step synthesis of nano-hydroxyapatite/graphene nanosheet hybrid coatings on Ti6Al4V alloys by hydrothermal method and their in-vitro corrosion responses. *Surf. Coat. Technol.* **2020**, *394*, 125858. [[CrossRef](#)]
38. Oskouei, R.H.; Fallahnezhad, K.; Kuppusami, S. An Investigation on the Wear Resistance and Fatigue Behaviour of Ti-6Al-4V Notched Members Coated with Hydroxyapatite Coatings. *Materials* **2016**, *9*, 111. [[CrossRef](#)] [[PubMed](#)]
39. Chen, S.H.; Ho, S.C.; Chang, C.H.; Chen, C.C.; Say, W.C. Influence of roughness on in-vivo properties of titanium implant surface and their electrochemical behavior. *Surf. Coat. Technol.* **2016**, *302*, 215–226. [[CrossRef](#)]
40. Sari, M.; Kristiano, N.A.; Ana, I.D.; Yusuf, Y. Carbonated Hydroxyapatite-Based Honeycomb Scaffold Coatings on a Titanium Alloy for Bone Implant Application—Physicochemical and Mechanical Properties Analysis. *Coatings* **2021**, *11*, 941. [[CrossRef](#)]
41. Kierzkowska, A.; Krasicka-Cydzik, E. Behaviour of Ti6Al4V implant alloy In Vitro after plastic deformation by bending. *Surf. Interface Anal.* **2008**, *40*, 507–512. [[CrossRef](#)]
42. Saruta, J.; Sato, N.; Ishijima, M.; Okubo, T.; Hirota, M.; Ogawa, T. Disproportionate Effect of Sub-Micron Topography on Osteoconductive Capability of Titanium. *Int. J. Mol. Sci.* **2019**, *20*, 4027. [[CrossRef](#)]
43. Cho, Y.-D.; Kim, W.-J.; Kim, S.; Ku, Y.; Ryoo, H.-M. Surface Topography of Titanium Affects Their Osteogenic Potential through DNA Methylation. *Int. J. Mol. Sci.* **2021**, *22*, 2406. [[CrossRef](#)]
44. Ball, M.; Grant, D.M.; Lo, W.-J.; Scotchford, C.A. The effect of different surface morphology and roughness on osteoblast-like cells. *J. Biomed. Mater. Res. A* **2007**, *86*, 637–647. [[CrossRef](#)] [[PubMed](#)]
45. ISO. 5832-3:2016. *Implants for Surgery—Metallic Materials—Part 3: Wrought Titanium 6-Aluminium 4-Vanadium Alloy*; ISO: Geneva, Switzerland, 2016.
46. Iqbal, A.; Zhao, G.; Zaini, J.; Gupta, M.K.; Jamil, M.; He, N.; Nauman, M.M.; Mikolajczyk, T.; Pimenov, D.Y. Between-the-Holes Cryogenic Cooling of the Tool in Hole-Making of Ti-6Al-4V and CFRP. *Materials* **2021**, *14*, 795. [[CrossRef](#)] [[PubMed](#)]
47. Singh, G.; Pruncu, C.I.; Gupta, M.K.; Mia, M.; Khan, A.M.; Jamil, M.; Pimenov, D.Y.; Sen, B.; Sharma, V.S. Investigations of Machining Characteristics in the Upgraded MQL-Assisted Turning of Pure Titanium Alloys Using Evolutionary Algorithms. *Materials* **2019**, *12*, 999. [[CrossRef](#)]
48. Mia, M.; Gupta, M.K.; Singh, G.; Królczyk, G.; Pimenov, D.Y. An approach to cleaner production for machining hardened steel using different cooling-lubrication conditions. *J. Clean. Prod.* **2018**, *187*, 1069–1081. [[CrossRef](#)]
49. Salur, E.; Kuntoğlu, M.; Aslan, A.; Pimenov, D.Y. The Effects of MQL and Dry Environments on Tool Wear, Cutting Temperature, and Power Consumption during End Milling of AISI 1040 Steel. *Metals* **2021**, *11*, 1674. [[CrossRef](#)]
50. Gupta, M.K.; Khan, A.M.; Song, Q.; Liu, Z.; Khalid, Q.S.; Jamil, M.; Kuntoğlu, M.; Usca, Ü.A.; Sarıkaya, M.; Pimenov, D.Y. A review on conventional and advanced minimum quantity lubrication approaches on performance measures of grinding process. *Int. J. Adv. Manuf. Syst.* **2021**, *117*, 729–750. [[CrossRef](#)]
51. Maruda, R.W.; Krolczyk, G.M.; Feldshtein, E.; Pusavec, F.; Szydłowski, M.; Legutko, S.; Sobczak-Kupiec, A. A study on droplets sizes, their distribution and heat exchange for minimum quantity cooling lubrication (MQCL). *Int. J. Mach. Tools Manuf.* **2016**, *100*, 81–92. [[CrossRef](#)]
52. ISO. ISO 5436-1:2000(en). *Geometrical Product Specifications (GPS)—Surface Texture: Profile Method—Terms, Definitions and Surface Texture Parameters*; ISO: Geneva, Switzerland, 2000.
53. Horvath, R.; Czifra, A.; Dregelyi-Kiss, A. Effect of conventional and non-conventional tool geometries to skewness and kurtosis of surface roughness in case of fine turning of aluminium alloys with diamond tools. *Int. J. Adv. Manuf. Technol.* **2014**, *78*, 297–304. [[CrossRef](#)]
54. Sánchez Egea, A.J.; Martynenko, V.; Simoncelli, A.; Serranoli, G.; Krahmer, D.M. Sliding abrasive wear when combining WEDM conditions and polishing treatment on H13 disks over 1045 carbon steel pins. *Int. J. Adv. Manuf. Technol.* **2021**. [[CrossRef](#)]
55. Kokubo, T.; Takadama, H. How useful is SBF in predicting In Vivo bone bioactivity? *Biomaterials* **2006**, *27*, 2907–2915. [[CrossRef](#)]
56. ISO. 16773-2:2016. *Electrochemical Impedance Spectroscopy (EIS) on Coated and Uncoated Metallic Specimens—Part 2: Collection of Data*; ISO: Geneva, Switzerland, 2016.
57. ISO. 16773-3:2016. *Electrochemical Impedance Spectroscopy (EIS) on Coated and Uncoated Metallic Specimens—Part 3: Processing and Analysis of Data from Dummy Cells*; ISO: Geneva, Switzerland, 2016.
58. Feldshtein, E.; Dyachkova, L.N.; Królczyk, G.M. On the evaluation of certain strength characteristics and fracture features of iron-based sintered MMCs with nanooxide additives. *Mater. Sci. Eng. A* **2019**, *756*, 455–463. [[CrossRef](#)]
59. Gacia-Martinez, E.; Miguel, V.; Martinez-Martinez, A.; Manjabacas, M.C.; Coello, J. Sustainable lubrication methods for the machining of titanium alloys: An overview. *Materials* **2019**, *12*, 3852. [[CrossRef](#)]
60. Alla, R.K.; Ginjupalli, K.; Upadhya, N.; Shamma, M.; Ravi, R.K.; Sekhar, R. Surface roughness of implant: A review. *Trends Biomater. Artif. Organs* **2011**, *25*, 112–118.
61. Krolczyk, G.M.; Maruda, R.W.; Krolczyk, J.B.; Nieslony, P.; Wojciechowski, S.; Legutko, S. Parametric and nonparametric description of the surface topography in the dry and MQCL cutting conditions. *Measurement* **2018**, *121*, 225–239. [[CrossRef](#)]
62. Leksycki, K.; Feldshtein, E.; Lisowicz, J.; Chudy, R.; Mrugalski, R. Cutting Forces and Chip Shaping When Finish Turning of 17-4 PH Stainless Steel under Dry, Wet, and MQL Machining Conditions. *Metals* **2020**, *10*, 1187. [[CrossRef](#)]
63. Leksycki, K.; Feldshtein, E.; Królczyk, G.M.; Legutko, S. On the Chip Shaping and Surface Topography When Finish Cutting 17-4 PH Precipitation-Hardening Stainless Steel under Near-Dry Cutting Conditions. *Materials* **2020**, *13*, 2188. [[CrossRef](#)] [[PubMed](#)]

64. Leksycki, K.; Feldshtein, E. The surface texture of Ti6Al4V titanium alloy under wet and dry finish turning conditions. In *Industrial Measurements in Machining, Lecture Notes in Mechanical Engineering*; Królczyk, G., Niesłony, P., Królczyk, J., Eds.; Springer: Cham, Switzerland, 2020; pp. 33–44. [[CrossRef](#)]
65. Stango, S.A.X.; Karthick, D.; Swaroop, S.; Mudali, U.K.; Vijayalakshimi, U. Development of hydroxyapatite coatings on laser textured 316 LSS and Ti-6Al-4V and its electrochemical behavior in SBF solution for orthopaedic applications. *Ceram. Int.* **2018**, *44*, 3149–3160. [[CrossRef](#)]
66. Krasicka-Cydzik, E. *Formowanie Cienkich warstw anodowych na tytanie i jego implantowych stopach w środowisku kwasu fosforowego*; Oficyna Wydawnicza Uniwersytetu Zielonogórskiego: Zielona Góra, Poland, 2003. (In Polish)